

Gas eller ånga som skall detekteras	Kemisk formel	Rör Diameter	Mätområde (ppm)	Antal pump drag	Time per Pump Stroke	Lagrings tid (År)	Nödvändig fukthalts korrigering: T=Temp. H=Humidity	Rör namn	Artikel nr	Antal Mätningar
Acetaldehyde	C2H4O	Thick	1 - 10	20	20 sec	2	-	Formaldehyde-0,1	D5086813	10
Acetaldehyde	C2H4O	Thick	5 - 50	10	20 sec	2	-	Formaldehyde-0,1	D5086813	10
Acetaldehyde	C2H4O	Thin	10 - 300	2	1,5 min	1	-	SAT-4M	UNI-CF000621	10
Acetic Acid	CH3COOH	Thick	1 - 15	5	20 sec	2	-	Acetic Acid-1	D5086821	10
Acetic Acid	CH3COOH	Thick	5 - 80	1	20 sec	2	-	Acetic Acid-1	D5086821	10
Acetic Acid	CH3COOH	Thick	n/a	n/a	n/a	2	-	Acetic Acid-1-ETW	D5093821	10
Acetic Acid	CH3COOH	Thick	2,5 - 50	1	1 min	1	-	SAA-2	UNI-CF000001	10
Acetone	C3H6O	Thick	100 - 1000	6	30 sec	2	-	Acetone-100	D5086829	10
Acetone	C3H6O	Thick	25 - 3500	3	40 sec	2	-	MEK-50	D5086837	10
Acetone	C3H6O	Thick	n/a	n/a	n/a	2	-	Acetone-100-ETW	D5093829	10
Acetone	C3H6O	Thin	50 - 4000	2	1,5 min	1	-	SAC-5	UNI-CF000002	10
Acetylene	C2H2	Thick	400 - 8300	1	20 sec	2	-	CO-0,1 %	D5085822	10
Acetylene	C2H2	Thin	100 - 5000	1	1 min	1	-	SAL-5	UNI-CF000003	10
Acid Gases	-	Thin	1 - 20	4	1,5 min	1	-	SAG-2	UNI-CF000680	10
Acid Gases	-	Thin	2 - 40	2	1,5 min	1	-	SAG-2	UNI-CF000680	10
Acid Gases	-	Thin	4 - 80	1	1,5 min	1	-	SAG-2	UNI-CF000680	10
Acrolein	C3H4O	Thick	0,2 - 10	20	20 sec	2	-	Acrolein-0,1-ETW	10017881	10
Acrolein	C3H4O	Thick	0,1 - 5	40	20 sec	2	-	Acrolein-0,1	10017882	10
Acrolein	C3H4O	Thick	0,2 - 10	20	20 sec	2	-	Acrolein-0,1	10017882	10
Acrolein	C3H4O	Thin	10 - 800	2	2 min	1	-	SAE-4	UNI-CF000745	10
Acrylonitrile	C3H3N	Thick	5 - 30	3	20 sec	2	-	Acrylonitrile-0,5	10016376	10
Acrylonitrile	C3H3N	Thick	1 - 10	10	20 sec	2	-	Acrylonitrile 0,5	10016376	10

Acrylonitrile	C3H3N	Thick	0,5 - 5	20	20 sec	2	-	Acrylonitrile-0,5	10016376	10
Acrylonitrile	C3H3N	Thick	0,5 - 5	20	20 sec	2	-	Acrylonitrile-0,5-ETW	10016607	5
Acrylonitrile	C3H3N	Thick	1 - 10	10	20 sec	2	-	Acrylonitrile 0,5-ETW	10016607	5
Acrylonitrile	C3H3N	Thick	5 - 30	3	20 sec	2	-	Acrylonitrile-0,5-ETW	10016607	5
Acrylonitrile	C3H3N	Thin	5 - 120	2	2 min	1	-	SAN-4L	UNI-CF000004	5
Adamsite (DM)	C12H9AsClN	Thick	>3 mg/m3	30	10 sec	2	-	CWA-DM-Tubes	7Sol-011	10
Alcohol	CxHxOH	Thick	100 - 6000	10	15 sec	2	-	Ethanol-100	D5086818	10
Alcohol	CxHxOH	Thin	0.05 - 5%	2	1,5 min	2	-	SET-7	UNI-CF000062	10
Amines	-	Thick	5 - 30	1	20 sec	2	-	Triethyl Amine-5	D5086816	10
Amines	-	Thin	2.5 - 50	2	1 min	1	-	SMA-3	UNI-CF000668	10
Amines	-	Thin	5 - 100	1	1 min	1	-	SMA-3	UNI-CF000668	10
Amines	-	Thin	10 - 200	½	1 min	1	-	SMA-3	UNI-CF000668	10
Ammonia	NH3	Thin	0,5 - 15	2	1 min	2	T + H	NH3 Ammonia	10-100-05	10
Ammonia	NH3	Thin	1 - 30	1	1 min	2	T + H	NH3 Ammonia	10-100-05	10
Ammonia	NH3	Thin	2 - 60	1	1 min	2	T + H	NH3 Ammonia	10-100-05	10
Ammonia	NH3	Thin	1 - 20	1	1 min	3	-	105SD	105SD	10
Ammonia	NH3	Thin	0,5 - 10	2	1 min	3	-	105SD	105SD	10
Ammonia	NH3	Thin	0,2 - 4	5	1 min	3	-	105SD	105SD	10
Ammonia	NH3	Thick	Warning Marks	2	35 sec	2	-	NH3-A	D5085755	10
Ammonia	NH3	Thick	20 - 1400	1	25 sec	2	-	NH3-20	D5085814	10
Ammonia	NH3	Thick	20 - 1400	1	25 sec	2	-	NH3-20	D5085914	200
Ammonia	NH3	Thick	0,1 - 10 %	2	25 sec	2	-	NH3-0,1%	D5085815	10
Ammonia	NH3	Thick	0,5 - 20 %	10	25 sec	2	-	NH3-0,1%	D5085815	10
Ammonia	NH3	Thick	2 - 80	10	25 sec	2	-	NH3-2	D5085845	10
Ammonia	NH3	Thick	100 - 600	2	25 sec	2	-	NH3-2	D5085845	10
Ammonia	NH3	Thick	100 - 600	2	25 sec	2	-	NH3-2	D5085945	200

Ammonia	NH3	Thick	0,25 - 3	5	20 sec	2	-	Triethyl Amine-5	D5086816	10
Ammonia	NH3	Thick	100 - 2000	1	20 sec	2	-	NH3-100	D5086859	10
Ammonia	NH3	Thick	n/a	n/a	n/a	2	-	NH3-2-ETW	D5092845	10
Ammonia	NH3	Thin	1.25 - 25	2	1 min	2	-	SAM-2	UNI-CF000005	10
Ammonia	NH3	Thin	2.5 - 50	1	1 min	2	-	SAM-2	UNI-CF000005	10
Ammonia	NH3	Thin	5 - 100	½	1 min	2	-	SAM-2	UNI-CF000005	10
Ammonia	NH3	Thin	2.5 - 50	2	30 sec	2	-	SAM-3	UNI-CF000006	10
Ammonia	NH3	Thin	5 -100	1	30 sec	2	-	SAM-3	UNI-CF000006	10
Ammonia	NH3	Thin	10 - 200	½	30 sec	2	-	SAM-3	UNI-CF000006	10
Ammonia	NH3	Thin	5 - 150	2	1 min	2	-	SAM-4M	UNI-CF000007	10
Ammonia	NH3	Thin	10-300	1	1 min	2	-	SAM-4M	UNI-CF000007	10
Ammonia	NH3	Thin	20 - 600	½	1 min	2	-	SAM-4M	UNI-CF000007	10
Ammonia	NH3	Thin	25 - 500	2	1 min	2	-	SAM-4	UNI-CF000008	10
Ammonia	NH3	Thin	50 -1000	1	1 min	2	-	SAM-4	UNI-CF000008	10
Ammonia	NH3	Thin	100 - 2000	½	1 min	2	-	SAM-4	UNI-CF000008	10
Ammonia	NH3	Thin	125 - 2500	2	1 min	2	-	SAM-5	UNI-CF000009	10
Ammonia	NH3	Thin	250 - 5000	1	1 min	2	-	SAM-5	UNI-CF000009	10
Ammonia	NH3	Thin	500 - 10.000	½	1 min	2	-	SAM-5	UNI-CF000009	10
Ammonia	NH3	Thin	0.25 - 5%	2	1 min	2	-	SAM-8M	UNI-CF000010	10
Ammonia	NH3	Thin	0.5 -10%	1	1 min	2	-	SAM-8M	UNI-CF000010	10
Ammonia	NH3	Thin	1 - 20%	½	1 min	2	-	SAM-8M	UNI-CF000010	10
Ammonia	NH3	Thin	0.5 - 15%	2	1 min	2	-	SAM-9	UNI-CF000011	10
Ammonia	NH3	Thin	1 - 30%	1	1 min	2	-	SAM-9	UNI-CF000011	10
Ammonia	NH3	Thin	2 - 60%	½	1 min	2	-	SAM-9	UNI-CF000011	10
Aniline		Thin	2 - 30	1	30 sec	1	-	SAI-2	UNI-CF000013	10
Arsine (Hydrogen Arsenide)	AsH3	Thick	0,05 - 1	10	1 min	2	-	Arsine-0,05	10072098	10
Arsine (Hydrogen Arsenide)	AsH3	Thin	0.5 - 15	2	1 min	2	-	SAR-2	UNI-CF000014	10
Arsine (Hydrogen Arsenide)	AsH3	Thin	1 - 30	1	1 min	2	-	SAR-2	UNI-CF000014	10
Arsine (Hydrogen Arsenide)	AsH3	Thin	2 - 60	½	1 min	2	-	SAR-2	UNI-CF000014	10

Arsine (Hydrogen Arsenide, SA)	AsH3	Thick	>1mg/m3	30	10 sec	2	-	CWA-SA.L.DA.DC-Tubes	7Sol-010	10
Aviation Oil		Thin	0.1 - 5 mg/l	1	1 min	1	-	SAO-4	UNI-CF000015	10
Benzene	C6H6	Thick	Warning Marks	n/a	n/a	2	-	Aromatic-A	D5085770	10
Benzene	C6H6	Thick	5 - 100	10	1 min	3	-	C6H6-5	D5085816	10
Benzene	C6H6	Thick	5 - 300	10	15 sec	2	-	Aromatic-HC	D5086811	10
Benzene	C6H6	Thick	1 - 25	15	1 min	2	-	C6H6-1	D5086835	10
Benzene	C6H6	Thick	5 - 100	4	1 min	2	-	C6H6-1	D5086835	10
Benzene	C6H6	Thick	0,25 - 5	15	50 sec	1	-	Benzene 0,2 (Specific)	D5086852	5
Benzene	C6H6	Thick	0,5 - 10	10	50 sec	1	-	Benzene 0,2 (Specific)	D5086852	5
Benzene	C6H6	Thick	n/a	n/a	n/a	3	-	C6H6-5-ETW	D5092816	10
Benzene	C6H6	Thin	0.25 - 12	5	2,5 min	2	-	SBE-2L	UNI-CF000016	5
Benzene	C6H6	Thin	1.25 - 25	4	2,5 min	2	-	SBE-2	UNI-CF000017	10
Benzene	C6H6	Thin	2.5 - 50	2	2,5 min	2	-	SBE-2	UNI-CF000017	10
Benzene	C6H6	Thin	5 - 100	1	2,5 min	2	-	SBE-2	UNI-CF000017	10
Benzene	C6H6	Thin	2.5 - 50	2	2 min	2	-	SBE-3	UNI-CF000018	10
Benzene	C6H6	Thin	5 - 100	1	2 min	2	-	SBE-3	UNI-CF000018	10
Benzene	C6H6	Thin	10 - 200	½	2 min	2	-	SBE-3	UNI-CF000018	10
Benzene (with toluene indicator)	C6H6	Thick	0 - 1	4	45 sec	2	-	Benzene 1/a	8103641	10
Bromine	Br2	Thick	0,2 - 3	10	20 sec	2	-	CL2-0,2	D5085801	10
Bromine	Br2	Thin	1 - 20	2	1 min	2	-	SBR-2M	UNI-CF000019	10
Bromine	Br2	Thin	5 - 100	1	1,5 min	2	-	SBR-3	UNI-CF000020	10
Bromobenzene	C6H5Br	Thick	30 - 720	10	15 sec	2	-	Aromatic-HC	D5086811	10
Bromoethane	C2H5Br	Thick	15 - 400	10	20 sec	2	-	Trichloroethane-5	D5086834	10
Bromoethane	C2H5Br	Thick	20 - 270	10	20 sec	2	-	Trichloroethane-5	D5086834	10
Bromoethane	C2H5Br	Thick	2 - 100	6	n/a	2	-	Methyl Bromide-2	D5086845	5
Bromoethane	C2H5Br	Thick	200 - 8000	2	20 sec	2	-	Methyl Bromide-200	D5086847	10

Bromoethane	C2H5Br	Thick	2000 - 20.000	1	20 sec	2	-	Methyl Bromide-200	D5086847	10
Bromopropane 1	C3H7Br	Thick	20 - 270	10	20 sec	2	-	Trichloroethane-5	D5086834	10
Butadiene, 1,3	C4H6	Thick	200 - 3800	5	20 sec	2	-	Propane-200	D5086831	10
Butadiene, 1,3	C4H6	Thick	100 - 1200	5	20 sec	2	-	Ethylene-50	D5086833	10
Butadiene, 1,3	C4H6	Thin	1 - 20	1	2 min	1	-	SBU-2M	UNI-CF000021	5
Butadiene, 1,3	C4H6	Thin	5 - 100	1	2 min	1	-	SBU-3	UNI-CF000022	5
Butane, n		Thick	200 - 4000	3	20 sec	2	-	Propane-200	D5086831	10
Butane, n		Thin	50 - 1400	1	4,5 min	1	-	SBT-5M	UNI-CF000627	10
Butanol, 1		Thin	10 - 150	2	2,5 min	1	-	SNB-4L	UNI-CF000612	10
Butanol, 2		Thin	10- 150	2	3,5 min	1	-	SIB-4L	UNI-CF000557	10
Butanol, 2		Thin	20 - 300	1	3,5 min	1	-	SIB-4L	UNI-CF000557	10
Butyl Acetate		Thin	0.005 - 1%	1	1,5 min	1	-	SBA-6	UNI-CF000023	10
Butyl Acrylate		Thin	5 - 60	3	2 min	1	-	SBC-3L	UNI-CF000024	10
Butyl Alcohol, iso (Isobutanol)	C4H10O	Thick	Warning Marks	15 sec	20 sec	2	-	Alcohols-A	D5085773	10
Butyl Alcohol, iso (Isobutanol)	C4H10O	Thick	150 - 2900	10	15 sec	2	-	Ethanol-100	D5086818	10
Butyl Alcohol, n (n-Butanol)	C4H10O	Thick	Warning Marks	15 sec	20 sec	2	-	Alcohols-A	D5085773	10
Butyl Alcohol, n (n-Butanol)	C4H10O	Thick	100 - 3900	10	15 sec	2	-	Ethanol-100	D5086818	10
Butyl Alcohol, s (s-Butanol)	C4H10O	Thick	Warning Marks	15 sec	20 sec	2	-	Alcohols-A	D5085773	10
Butyl Alcohol, s (s-Butanol)	C4H10O	Thick	300 - 1500	10	15 sec	2	-	Ethanol-100	D5086818	10
Butyl Amine	C4H11N	Thick	3 - 36	1	20 sec	2	-	Triethyl Amine-5	D5086816	10
Butyl Amine, n	C4H11N	Thin	5 - 100	1	1 min	1	-	SMA-3	UNI-CF000668	10
Butyl Amine, s	C4H11N	Thick	2 - 18	1	20 sec	2	-	Triethyl Amine-5	D5086816	10
Butyl Amine, t	C4H11N	Thick	2 - 14	1	20 sec	2	-	Triethyl Amine-5	D5086816	10
Butyl Mercaptan, n	C4H10S	Thick	1,5 - 15	20	30 sec	2	-	Ethyl Mercaptan-0,5	D5086815	10

Butyl Mercaptan, t	C4H10S	Thick	0,8 - 5	15	30 sec	2	-	Ethyl Mercaptan-0,5	D5086815	10
Butyl Mercaptan, t		Thin	2.5 - 60	2	1,5 min	1	-	SBM-4L	UNI-CF000060	10
Butyl Mercaptan, t		Thin	5 - 120	1	1,5 min	1	-	SBM-4L	UNI-CF000060	10
Butyl Mercaptan, t		Thin	10 - 240	½	1,5 min	1	-	SBM-4L	UNI-CF000060	10
Butyl Mercaptan, t		Thin	12.5 - 250	2	1,5 min	1	-	SEM-4M	UNI-CF000061	10
Butyl Mercaptan, t		Thin	25 - 500	1	1,5 min	1	-	SEM-4M	UNI-CF000061	10
Butyl Mercaptan, t		Thin	0.5 - 15 mg/m3 (0.1 4 ppm)	3	1,5 min	1	-	SBM-1M	UNI-CF000694	10
Butyl Mercaptan, t		Thin	2 - 80	1	1,5 min	1	-	SEM-(3L)	UNI-CF000784	10
Butyl Mercaptan, t		Thin	0.5 - 30 mg/m3 (0.1 8 ppm)	2	1 min	1	-	SBM-1	UNI-CF00138	10
Butylene, 1 (Olefins)	C4H8	Thick	100 - 5000	10	20 sec	2	-	Ethylene-50	D5086833	10
Butylene, 1 (Olefins)	C4H8	Thick	Warning Marks	2	20 sec	2	-	Gasoline-A	D5085769	10
Butylene, 2 (Olefins)	C4H8	Thick	200 - 5000	10	20 sec	2	-	Ethylene-50	D5086833	10
Butylene, 2 (Olefins)	C4H8	Thick	Warning Marks	2	20 sec	2	-	Gasoline-A	D5085769	10
BZ (3-Quinuclidinyl benzilate)	C21H23NO 3	Thick	0,5 - 100 mg/m3	20	10 sec	2	-	CWA-BZ-Tubes	7Sol-009	10
BZ (3-Quinuclidinyl benzilate)	C21H23NO 3	Thick	1 mg/m3	30	10 sec	2	-	BZ-CWA	D5086885	10
Carbon Dioxide	CO2	Thick	5 - 40 %	n/a	n/a	#####	-	CO2 in Biogas	10012107	10
Carbon Dioxide	CO2	Thick	Warning Marks	2	20 sec	2	-	CO2-A	D5085758	10
Carbon Dioxide	CO2	Thick	0,1 - 7 %	1	20 sec	2	-	CO2-0,1%	D5085817	10
Carbon Dioxide	CO2	Thick	1 - 20 %	1	20 sec	2	-	CO2-1%	D5085841	10
Carbon Dioxide	CO2	Thick	100 - 2000	1,5 liter	5 min	2	-	CO2-HP	D5085848	10
Carbon Dioxide	CO2	Thick	0,1 - 7 %	1	20 sec	2	-	CO2-0,1%	D5085917	200
Carbon Dioxide	CO2	Thick	1 - 20 %	1	20 sec	2	-	CO2-1%	D5085941	200

Carbon Dioxide	CO2	Thick	100 - 3000	10	30 sec	2	-	CO2-100	D5086814	10
Carbon Dioxide	CO2	Thick	0,5 - 10 %	2	30 sec	2	-	CO2-0,5%	D5086844	10
Carbon Dioxide	CO2	Thick	0,5 - 10 %	2	30 sec	2	-	CO2-0,5%	D5086944	200
Carbon Dioxide	CO2	Thick	n/a	n/a	n/a	2	-	CO2-0,1%-ETW	D5092817	10
Carbon Dioxide	CO2	Thin	150 - 2500	4	3 min	2	T	SCD-5	UNI-CF000025	10
Carbon Dioxide	CO2	Thin	300 - 5000	2	3 min	2	T	SCD-5	UNI-CF000025	10
Carbon Dioxide	CO2	Thin	0.06 - 1.0%	1	3 min	2	T	SCD-5	UNI-CF000025	10
Carbon Dioxide	CO2	Thin	0.125 - 1.5%	2	2,5 min	2	T	SCD-7	UNI-CF000027	10
Carbon Dioxide	CO2	Thin	0.25 - 3%	1	2,5 min	2	T	SCD-7	UNI-CF000027	10
Carbon Dioxide	CO2	Thin	0.5 - 6%	½	2,5 min	2	T	SCD-7	UNI-CF000027	10
Carbon Dioxide	CO2	Thin	0.25 - 5%	2	2,5 min	2	T	SCD-8M	UNI-CF000028	10
Carbon Dioxide	CO2	Thin	0.5 - 10%	1	2,5 min	2	T	SCD-8M	UNI-CF000028	10
Carbon Dioxide	CO2	Thin	1 - 20%	½	2,5 min	2	T	SCD-8M	UNI-CF000028	10
Carbon Dioxide	CO2	Thin	0.5 - 10%	2	2,5 min	2	T	SCD-8	UNI-CF000029	10
Carbon Dioxide	CO2	Thin	1 - 20%	1	2,5 min	2	T	SCD-8	UNI-CF000029	10
Carbon Dioxide	CO2	Thin	2 - 40%	½	2,5 min	2	T	SCD-8	UNI-CF000029	10
Carbon Dioxide	CO2	Thin	2.5 - 30%	1	4 min	2	T	SCD-10	UNI-CF000030	10
Carbon Dioxide	CO2	Thin	5 - 60%	½	4 min	2	T	SCD-10	UNI-CF000030	10
Carbon Disulfide	CS2	Thick	2 - 300	1	20 sec	2	-	CS2-2	D5085834	10
Carbon Disulfide	CS2	Thick	2 - 300	1	20 sec	2	-	CS2-2	D5085934	200
Carbon Disulfide	CS2	Thick	n/a	n/a	n/a	2	-	CS2-2-ETW	D5092834	10
Carbon Disulfide	CS2	Thin	1.25 - 20	4	1,5 min	2	-	SCS-2	UNI-CF000032	5
Carbon Disulfide	CS2	Thin	2.5 - 50	2	1,5 min	2	-	SCS-2	UNI-CF000032	5
Carbon Disulfide	CS2	Thin	5 - 100	½	1,5 min	2	-	SCS-2	UNI-CF000032	5
Carbon Disulfide	CS2	Thin	2.5 - 50	2	1,5 min	2	-	SCS-3	UNI-CF000033	5
Carbon Disulfide	CS2	Thin	5 - 100	1	1,5 min	2	-	SCS-3	UNI-CF000033	5
Carbon Disulfide	CS2	Thin	10 - 200	½	1,5 min	2	-	SCS-3	UNI-CF000033	5
Carbon Disulfide	CS2	Thin	50 - 1600	1	1,5 min	2	-	SCS-5M	UNI-CF000034	5
Carbon Monoxide	CO	Thick	Warning Marks	2	20 sec	2	-	CO-A	D5085759	10
Carbon Monoxide	CO	Thick	0,01 - 0,3 %	1	20 sec	3	-	CO-10Vol%	D5085820	10
Carbon Monoxide	CO	Thick	10 - 300	10	20 sec	3	-	CO-10Vol%	D5085820	10
Carbon Monoxide	CO	Thick	10 - 300	10	20 sec	3	-	CO-10ppm	D5085821	10

Carbon Monoxide	CO	Thick	100 - 3000	1	20 sec	3	-	CO-10ppm	D5085821	10
Carbon Monoxide	CO	Thick	0,1 - 1,0 %	1	20 sec	2	-	CO-0,1%	D5085822	10
Carbon Monoxide	CO	Thick	0,3 - 7 %	1	20 sec	2	-	CO-0,5%	D5085835	10
Carbon Monoxide	CO	Thick	5 - 1000	10	20 sec	2	-	CO-5	D5085836	10
Carbon Monoxide	CO	Thick	5 - 70	1,5 liter	5 min	2	-	CO-HP	D5085847	10
Carbon Monoxide	CO	Thick	10 - 300	10	20 sec	3	-	CO-10Vol%	D5085920	200
Carbon Monoxide	CO	Thick	100 - 3000	1	20 sec	3	-	CO-10ppm	D5085921	200
Carbon Monoxide	CO	Thick	5 - 1000	10	20 sec	2	-	CO-5	D5085936	200
Carbon Monoxide	CO	Thick	3000 - 70.000	1	20 sec	2	-	CO-3000	D5086839	10
Carbon Monoxide	CO	Thick	n/a	n/a	n/a	3	-	CO-10ppm-ETW	D5092821	10
Carbon Monoxide	CO	Thin	2.5 - 50	4	3 min	1	-	SCO-3	UNI-CF000035	10
Carbon Monoxide	CO	Thin	5 - 100	2	3 min	1	-	SCO-3	UNI-CF000035	10
Carbon Monoxide	CO	Thin	10 - 200	1	3 min	1	-	SCO-3	UNI-CF000035	10
Carbon Monoxide	CO	Thin	12.5 - 300	2	4 min	2	-	SCO-4M	UNI-CF000037	10
Carbon Monoxide	CO	Thin	25 - 600	1	4 min	2	-	SCO-4M	UNI-CF000037	10
Carbon Monoxide	CO	Thin	50 - 1200	½	4 min	2	-	SCO-4M	UNI-CF000037	10
Carbon Monoxide	CO	Thin	25 - 500	2	1,5 min	2	-	SCO-4	UNI-CF000038	10
Carbon Monoxide	CO	Thin	50 - 1000	1	1,5 min	2	-	SCO-4	UNI-CF000038	10
Carbon Monoxide	CO	Thin	100 - 2000	½	1,5 min	2	-	SCO-4	UNI-CF000038	10
Carbon Monoxide	CO	Thin	50 - 1500	2	2,5 min	1	-	SCO-5	UNI-CF000039	10
Carbon Monoxide	CO	Thin	100 - 3000	1	2,5 min	1	-	SCO-5	UNI-CF000039	10
Carbon Monoxide	CO	Thin	200 - 6000	½	2,5 min	1	-	SCO-5	UNI-CF000039	10
Carbon Monoxide	CO	Thin	0.01 - 0.7%	1	3,5 min	2	-	SCO-6M	UNI-CF000040	10
Carbon Monoxide	CO	Thin	0.1 - 7%	1	3 min	2	-	SCO-8L	UNI-CF000041	10
Carbon Monoxide	CO	Thin	0.5 - 10%	2	4 min	1	-	SCO-8	UNI-CF000042	10
Carbon Monoxide	CO	Thin	1 - 20%	1	4 min	1	-	SCO-8	UNI-CF000042	10
Carbon Monoxide	CO	Thin	1 - 30	5	3 min	1	-	SCO-2	UNI-CF000753	10
Carbon Monoxide in H2	CO	Thin	5 - 100	3	1,5 min	1	-	SCH-3	UNI-CF000036	10
Carbon Monoxide in H2	CO	Thin	100 - 3000	1	2 min	1	-	SCH-5	UNI-CF000642	10

Carbon Tetrachloride (Tetrachloromethane)		Thin	1 - 60	1	2,5 min	1	-	SCT-3L	UNI-CF000043	5
Carbonyl Sulfide		Thin	2.5 - 50	4	1 min	2	-	SCU-3	UNI-CF000044	5
Carbonyl Sulfide		Thin	5 - 100	2	1 min	2	-	SCU-3	UNI-CF000044	10
Carbonyl Sulfide		Thin	10 - 200	½	1 min	2	-	SCU-3	UNI-CF000044	10
Chloracetophenon (Phenacyl Chloride, MACE, CN)	C8H7ClO	Thick	0,1 - 100 mg/m3	30	10 sec	2	-	CWA-CN-Tubes	7Sol-006	10
Chlorine	Cl2	Thick	Warning Marks	2	20 sec	2	-	CL2-A	D5085757	10
Chlorine	Cl2	Thick	0,2 - 3	10	20 sec	2	-	CL2-0,2	D5085801	10
Chlorine	Cl2	Thick	1 - 9	5	25 sec	2	-	CLO2-0,05	D5086812	10
Chlorine	Cl2	Thick	4 - 46	1	25 sec	2	-	CLO2-0,05	D5086812	10
Chlorine	Cl2	Thick	50 - 500	1	20 sec	2	-	CL2-50	D5086855	10
Chlorine	Cl2	Thick	n/a	n/a	n/a	2	-	CL0,2-ETW	D5092801	10
Chlorine	Cl2	Thin	0.25 - 5	2	1,5 min	2	-	SCL-1M	UNI-CF000045	10
Chlorine	Cl2	Thin	0.5 - 10	1	1,5 min	2	-	SCL-1M	UNI-CF000045	10
Chlorine	Cl2	Thin	1.25 - 25	2	1,5 min	2	-	SCL-2	UNI-CF000046	10
Chlorine	Cl2	Thin	2.5 - 50	1	1,5 min	2	-	SCL-2	UNI-CF000046	10
Chlorine	Cl2	Thin	5 - 100	½	1,5 min	2	-	SCL-2	UNI-CF000046	10
Chlorine	Cl2	Thin	2.5 - 50	2	1 min	2	-	SCL-3	UNI-CF000047	10
Chlorine	Cl2	Thin	5 - 100	1	1 min	2	-	SCL-3	UNI-CF000047	10
Chlorine	Cl2	Thin	10 - 200	½	1 min	2	-	SCL-3	UNI-CF000047	10
Chlorine	Cl2	Thin	25 - 250	2	2,5 min	2	-	SCL-4M	UNI-CF000048	10
Chlorine	Cl2	Thin	50 - 500	1	2,5 min	2	-	SCL-4M	UNI-CF000048	10
Chlorine	Cl2	Thin	100 -1000	½	2,5 min	2	-	SCL-4M	UNI-CF000048	10
Chlorine Dioxide	ClO2	Thick	0,25 - 15	1	25 sec	2	-	CLO2-0,05	D5086812	10
Chlorine Dioxide	ClO2	Thick	0,05 - 1,5	5	25 sec	2	-	CLO2-0,05	D5086812	10
Chlorine Dioxide	ClO2	Thin	0.05 - 2.5	2	2 min	1	-	SCI-1M	UNI-CF000706	10
Chlorine Dioxide	ClO2	Thin	0.1 - 5	1	2 min	1	-	SCI-1M	UNI-CF000706	10
Chlorine Dioxide	ClO2	Thin	0.2 - 10	½	2 min	1	-	SCI-1M	UNI-CF000706	10
Chlorine Dioxide	ClO2	Thin	100 - 4000	1	2,5 min	2	-	SCI-5	UNI-CF000796	10

Chlorobenzal malononitrile, -2 (Tear gas, CS)	C10H5ClN2	Thick	0,5 - 50 mg/m3	10	10 sec	2	-	CWA-CS-Tubes	7Sol-007	10
Chlorobenzene	C6H5Cl	Thick	Warning Marks	n/a	n/a	2	-	CKW-A	D5085771	10
Chlorobenzene	C6H5Cl	Thick	40 - 610	10	15 sec	2	-	Aromatic-HC	D5086811	10
Chlorobenzene	C6H5Cl	Thick	10 - 140	3	50 sec	2	-	Chlorobenzene-10-ETW	D5093811	10
Chlorobenzene	C6H5Cl	Thin	10 - 200	2	1,5 min	2	-	SCB-4L	UNI-CF000049	10
Chlorobromomethane	CH2BrCl	Thick	5 - 180	10	20 sec	2	-	Trichloroethane-5	D5086834	10
Chlorobutane, 1	C4H9Cl	Thick	5 - 170	10	20 sec	2	-	Trichloroethane-5	D5086834	10
Chloroethane	C2H5Cl	Thick	50 - 800	10	20 sec	2	-	Trichloroethane-5	D5086834	10
Chloroform (Trichloromethane)	C5H11Cl	Thick	8 - 100	10	20 sec	2	-	Trichloroethane-5	D5086834	10
Chloroform (Trichloromethane)		Thin	10 -100	5	2 min	1	-	SCF-3	UNI-CF000790	10
Chloropentane, 1	C5H11Cl	Thick	5 - 550	10	20 sec	2	-	Trichloroethane-5	D5086834	10
Chloropentane, 2	C5H11Cl	Thick	8 - 1700	10	20 sec	2	-	Trichloroethane-5	D5086834	10
Chloropicrin	CCl3NO2	Thin	0.013 - 2.1	5	2 min	1	-	SCP-2M	UNI-CF000718	5
Chloropicrin	CCl3NO2	Thin	0.04 - 6.4	2	2 min	1	-	SCP-2M	UNI-CF000718	5
Chloropicrin	CCl3NO2	Thin	0.1 - 16	1	2 min	1	-	SCP-2M	UNI-CF000718	5
Chloroprene	C4H5Cl	Thick	5 - 60	3	1 min	3	-	Chloroprene	6718901	10
Chloropropane, 1		Thick	5 - 220	10	20 sec	2	-	Trichloroethane-5	D5086834	10
Cresol, o (Orthocresol)	C7H8O	Thin	1 - 25	2	1,5 min	1	-	SCR-2M	UNI-CF000633	10
Cyanogen Chloride (CK)	CNCl	Thick	n/a	n/a	n/a	2	-	CK, AC, DP, CG	D5093884	10
Cyanogen Chloride (CK)	NCCl	Thick	1 - 500 mg/m3	15	10 sec	2	-	CWA-CK.AC.DP.CG-Tubes	7Sol-004	10
Cyanogen Chloride (CK)	NCCl	Thick	5 mg/m3	30	10 sec	2	-	CK, AC, DP, CG-CWA	D5086884	10

Cycloheptane	C7H14	Thick	80 - 3300	5	20 sec	2	-	Hexane-20	D5086832	10
Cyclohexane	C6H12	Thick	Warning Marks	2	20 sec	2	-	Gasoline-A	D5085769	10
Cyclohexane	C6H12	Thick	80 - 3400	2	20 sec	2	-	Hexane-20	D5086832	10
Cyclohexane	C6H12	Thick	20 - 1000	5	20 sec	2	-	Hexane-20	D5086832	10
Cyclohexane	C6H12	Thin	14 - 1680	1	1,5 min	1	-	SHE-5L	UNI-CF000068	10
Cyclohexane	C6H12	Thin	375 - 7800	1	1 min	2	-	SHE-6L	UNI-CF000069	10
Cyclohexanone	C6H10O	Thin	2 - 100	3	1,5 min	1	-	SCN-3	UNI-CF000052	10
Cyclohexyl Amine	C6H13N	Thick	7 - 38	2	20 sec	2	-	Triethyl Amine-5	D5086816	10
Cyclohexyl Amine	C6H13N	Thin	0.5 - 30	2	1,5 min	1	-	SMA-3	UNI-CF000668	10
Cyclooctane	C8H16	Thick	20 - 2100	10	20 sec	2	-	Hexane-20	D5086832	10
Cyclopentane	C5H10	Thick	80 - 2700	2	20 sec	2	-	Hexane-20	D5086832	10
Cyclosarin (GF)	C5H11N2O 2P	Thick	0,0005 - 10 mg/m3	50	10 sec	1	-	CWA- GB.GD.GP.GA.G F.VX-Tubes	7Sol-008	10
Cyclosarin (GF)	C5H11N2O 2P	Thick	0,005 - 100 mg/m3	5	10 sec	1	-	CWA- GB.GD.GP.GA.G F.VX-Tubes	7Sol-008	10
Cyclosarin (GF)	C5H11N2O 2P	Thick	>0,01 mg/m3	10	10 sec	1	-	Phosph. Acid Esters-CWA	D5086880	10
Decane, n	C10H22	Thick	50 - 500	10	20 sec	2	-	Hexane-20	D5086832	10
Dibenzoxazepine (CR)	C13H9NO	Thick	>0,1 mg/m3	30	10 sec	2	-	CWA-CR-Tubes	7Sol-012	10
Dibromoethane	C2H4Br2	Thick	9 - 200	10	20 sec	2	-	Trichloroethan e-5	D5086834	10
Dibromoethane, 1,1		Thick	25 - 700	10	20 sec	2	-	Trichloroethan e-5	D5086834	10
Dibromoethane, 1,2	C2H4Br2	Thick	25 - 700	10	20 sec	2	-	Trichloroethan e-5	D5086834	10
Dibromomethane, 1,2		Thick	25 - 700	10	20 sec	2	-	Trichloroethan e-5	D5086834	10
Dichlorobenzene, 1,2	C6H4Cl2	Thick	Warning Marks	n/a	n/a	2	-	CKW-A	D5085771	10
Dichlorobenzene, 1,2	C6H4Cl2	Thick	Warning Marks	n/a	n/a	2	-	CKW-A	D5085771	10

Dichlorobenzene, 1,2	C6H4Cl2	Thin	10 - 300	2	1,5 min	1	-	SDB-4M	UNI-CF000709	10
Dichloroethane, 1,1		Thick	8 - 300	10	20 sec	2	-	Trichloroethane-5	D5086834	10
Dichloroethane, 1,2		Thin	1 - 50	2	2,5 min	1	-	SDC-2	UNI-CF000653	5
Dichloroethylene, 1,1		Thick	10 - 600	10	20 sec	2	-	Trichloroethane-5	D5086834	10
Dichloroethylene, 1,1		Thick	10 - 500	10	20 sec	2	-	Trichloroethane-5	D5086834	10
Dichloroethylene, 1,2		Thick	10 - 500	10	20 sec	2	-	Trichloroethane-5	D5086834	10
Dichloroethylene, 1,2		Thin	20 - 400	1	2 min	1	-	SDH-4M	UNI-CF000742	5
Dichloromethane	CH2Cl2	Thick	Warning Marks	5	20 sec	2	-	Dichloromethane-A	D5085772	10
Dichloromethane	CH2Cl2	Thick	12,50 - 200	40	20 sec	2	-	Dichloromethane-50	D5085823	10
Dichloromethane	CH2Cl2	Thick	25 - 400	20	20 sec	2	-	Dichloromethane-50	D5085823	10
Dichloromethane	CH2Cl2	Thick	100 - 1000	10	20 sec	2	-	Dichloromethane-50	D5085823	10
Dichloropropane, 1,2	C3H6Cl2	Thick	5 - 440	10	20 sec	2	-	Trichloroethane-5	D5086834	10
Dichloropropane, 1,3	C3H6Cl2	Thick	5 - 220	10	20 sec	2	-	Trichloroethane-5	D5086834	10
Diesel		Thin	0.1 - 5 mg/l	1	1,5 min	2	-	SDL-4	UNI-CF000054	10
Diethyl Amine	C4H11N	Thick	3 - 27	2	20 sec	2	-	Triethyl Amine-5	D5086816	10
Diethyl Amine	C4H11N	Thin	2 - 20	2	1,5 min	1	-	SMA-3	UNI-CF000668	10
Diethyl Ether	C4H10O	Thin	0.04- 1%	1	2 min	2	-	SDE-6	UNI-CF000056	10
Dimethyl Amine	C2H7N	Thick	3 - 27	2	20 sec	2	-	Triethyl Amine-5	D5086816	10
Dimethyl Formamide (DMF)	C3H7NO	Thin	2 - 30	2	1,5 min	2	-	SMF-2	UNI-CF000053	5
Dimethyl Hydrazine, 1,1		Thick	0,05 - 1,1	5	50 sec	2	-	Hydrazine-0,1	D5086861	10

Dimethyl Sulfate	C2H6O4S	Thick	0,005 - 0,05	200	15 sec	3	-	Dimethyl Sulfate 0,005/c	6718701	9
Dimethyl Sulfide	C2H6S	Thin	0.5 - 10	1	2 min	1	-	SDS-2M	?	10
Dimethylbutane, 2,2		Thick	100 - 4900	2	20 sec	2	-	Hexane-20	D5086832	10
Dioxane, 1,4	C4H8O2	Thin	0,07 - 4%	1	2 min	2	-	SEO-7	UNI-CF000624	10
Diphenylchloroarsine (DA)	C12H10AsCl	Thick	>1mg/m3	30	10 sec	2	-	CWA-SA.L.DA.DC-Tubes	7Sol-010	10
Diphenylcyanoarsine (DC)	C13H10AsN	Thick	>1mg/m3	30	10 sec	2	-	CWA-SA.L.DA.DC-Tubes	7Sol-010	10
Diphosgene (DP)	C2Cl4O2	Thick	1 - 2000 mg/m3	15	10 sec	2	-	CWA-CK.AC.DP.CG-Tubes	7Sol-004	10
Diphosgene (DP)	C2Cl4O2	Thick	5 mg/m3	30	10 sec	2	-	CK, AC, DP, CG-CWA	D5086884	10
Epichlorohydrin	C3H5ClO	Thick	5 - 80	20	25 sec	3	-	Epichloro hydrin 5/c	6728111	10
Ethanol (Ethyl Alcohol)	C2H6O	Thick	100 - 3000	10	15 sec	2	-	Ethanol-100	D5086818	10
Ethanol (Ethyl Alcohol)	C2H6O	Thick	200 - 6000	5	15 sec	2	-	Ethanol-100	D5086818	10
Ethanol (Ethyl Alcohol)	C2H6O	Thick	n/a	n/a	n/a	2	-	Ethanol-100-ETW	D5093818	10
Ethanol (Ethyl Alcohol)	C2H6O	Thin	0.05 - 5%	2	1,5 min	2	-	SET-7	UNI-CF000062	10
Ethanol Amine	C2H7NO	Thin	18.5 - 370	1	1 min	1	-	SMA-3	UNI-CF000668	10
Ethyl Acetate	C4H8O2	Thick	200 - 3000	20	15 sec	2	T	Ethyl Acetate-200	10024253	10
Ethyl Acetate	C4H8O3	Thin	0.1 - 5%	1	2,5 min	2	-	SEA-7	UNI-CF000063	10
Ethyl Alcohol (Ethanol)	C2H6O	Thick	Warning Marks	15 sec	20 sec	2	-	Alcohols-A	D5085773	10
Ethyl Amine	C2H7N	Thick	4 - 55	1	20 sec	2	-	Triethyl Amine-5	D5086816	10
Ethyl Benzene	C8H10	Thick	Warning Marks	n/a	n/a	2	-	Aromaten-A	D5085770	10

Ethyl Benzene	C8H10	Thick	10 - 1800	5	40 sec	3	-	Tol-5	D5085828	10
Ethyl Benzene	C8H10	Thick	5 - 600	10	40 sec	3	-	Tol-5	D5085828	10
Ethyl Benzene	C8H10	Thin	5 - 150	1	1,5 min	2	-	SEB-4L	UNI-CF000057	10
Ethyl Benzene	C8H10	Thin	30 - 400	1	1,5 min	2	-	SEB-4M	UNI-CF000058	10
Ethyl Formate		Thin	20 - 500	1	3 min	1	-	SEF-4M	UNI-CF000630	10
Ethyl Glycol Acetate		Thick	50 - 700	10	20 sec	3	-	Ethyl Glycol Acetate 50/a	6726801	10
Ethyl Mercaptan		Thick	10 - 100	1	25 sec	2	-	Ethyl Mercaptan-0,5	D5086815	10
Ethyl Mercaptan		Thick	0,5 - 5	15	25 sec	2	-	Ethyl Mercaptan-0,5	D5086815	10
Ethyl Mercaptan		Thin	0.5 - 10	2	2 min	1	-	SEM-2M	UNI-CF000059	10
Ethyl Mercaptan		Thin	0,5 - 10	1	2 min	1	-	SEM-2M	UNI-CF000059	10
Ethyl Mercaptan		Thin	1 - 20	½	2 min	1	-	SEM-2M	UNI-CF000059	10
Ethyl Mercaptan		Thin	2.5 - 60	2	1,5 min	1	-	SEM-4L	UNI-CF000060	10
Ethyl Mercaptan		Thin	5 - 120	1	1,5 min	1	-	SEM-4L	UNI-CF000060	10
Ethyl Mercaptan		Thin	10 - 240	½	1,5 min	1	-	SEM-4L	UNI-CF000060	10
Ethyl Mercaptan		Thin	12.5 - 250	2	1,5 min	1	-	SEM-4M	UNI-CF000061	10
Ethyl Mercaptan		Thin	25 - 500	1	1,5 min	1	-	SEM-4M	UNI-CF000061	10
Ethyl Mercaptan		Thin	0.5 - 30 mg/m3 (0.1 8 ppm)	2	1 min	1	-	SBM-1	UNI-CF000138	10
Ethyl Mercaptan		Thin	0.5 - 15 mg/m3 (0.1 4 ppm)	3	1,5 min	1	-	SBM-1M	UNI-CF000694	10
Ethyl Mercaptan		Thin	2 - 40	2	1 min	1	-	SEM-(3L)	UNI-CF000784	10
Ethylene	C2H4	Thin	0,1 - 20	5	2 min	3	-	108B	108B	10
Ethylene	C2H4	Thin	0,5 - 100	1	2 min	3	-	108B	108B	10
Ethylene	C2H4	Thick	Warning Marks	2	20 sec	2	-	Gasoline-A	D5085769	10
Ethylene	C2H4	Thick	25 - 5000	5	20 sec	2	-	Ethylene-50	D5086833	10
Ethylene Diamine		Thick	5 - 27	2	20 sec	2	-	Triethyl Amine-5	D5086816	10

Ethylene Dibromide		Thin	0,5 - 25	2	2 min	1	-	SED-2	UNI-CF000668	5
Ethylene Dibromide		Thin	1 - 50	1	2 min	1	-	SED-2	UNI-CF000688	10
Ethylene Dibromide		Thin	2 - 100	½	2 min	1	-	SED-2	UNI-CF000688	10
Ethylene Glycol		Thin	10 - 100 mg/m3	2	2 min	1	-	SEG-2	UNI-CF000757	10
Ethylene Oxide		Thin	0.025 - 1.5%	2	2 min	2	-	SEO-7	UNI-CF000624	10
Ethylene Oxide		Thin	0.05 - 3%	1	2 min	2	-	SEO-7	UNI-CF000624	10
Ethylene Oxide		Thin	0.1 - 6%	½	2 min	2	-	SEO-7	UNI-CF000624	10
Ethylene Oxide		Thin	1 - 100	2	2 min	2	-	SEO-3	UNI-CF000748	10
Ethylene Oxide		Thin	0.1 - 10	4	1,5 min	1	-	SEO-1	UNI-CF000776	5
EX-Test (LEL detector tube)	-	Thick	0 - 10% LEL	3	15 sec	2	-	LEL-Tubes	7Sol-001	10
Extraction Gasolines		Thick	Qualitative	-	20 sec	2	-	BNZ-Test	D5085807	10
Fluorine	F2	Thick	0,05 - 1	40	15 sec	3	-	Fluorine 0,1/a	8101491	10
Fluorine	F2	Thick	0,1 - 2	20	15 sec	3	-	Fluorine 0,1/a	8101491	10
Formaldehyde	HCHNO	Thick	0,1 - 1	20	20 sec	2	-	Formaldehyde- 0,1	D5086813	10
Formaldehyde	HCHNO	Thick	1 - 10	10	20 sec	2	-	Formaldehyde- 0,1	D5086813	10
Formaldehyde	HCHNO	Thick	n/a	n/a	n/a	2	-	Formaldehyde- 0,1-ETW	D5093813	10
Formaldehyde	HCHNO	Thin	0.05 - 2	10	1 min	1	-	171SC	171SC	10
Formaldehyde	HCHNO	Thin	0.1 - 4	5	1 min	1	-	171SC	171SC	10
Formaldehyde	HCHNO	Thin	5 - 100	2	2,5 min	1	-	SFO-3	UNI-CF000065	10
Formaldehyde	HCHNO	Thin	0.1 - 5	5	2 min	1	-	SFO-1M	UNI-CF000735	10
Formic Acid		Thick	2 - 24	5	20 sec	2	-	Acetic Acid-1	D5086821	10
Formic Acid		Thick	10 - 160	1	20 sec	2	-	Acetic Acid-1	D5086821	10
Formic Acid		Thin	1 - 15	3	1,5 min	1	-	SFA-2M	UNI-CF000066	10
Furan		Thin	0.01 - 1%	1	2 min	1	-	SFU-6	UNI-CF000665	10
Furfural		Thin	1 - 16	4	2 min	2	-	SFL-2M	UNI-CF000067	10

Fururyl Alcohol		Thick	280 - 1500	10	15 sec	2	-	Phenol-1	D5086838	10
Fururyl Alcohol		Thick	170 - 1350	20	15 sec	2	-	Phenol-1	D5086838	10
Gasoline (Naphthalene)		Thin	10 - 100	1	1,5 min	1	-	153U	153U	10
Gasoline (Naphthalene)		Thick	Warning Marks (100- 2000)	2 to 15	20 sec	2	-	BNZ-Test	D5085807	10
Gasoline (Naphthalene)		Thick	30 - 600	10	20 sec	2	-	Gasoline-30	D5085898	10
Gasoline (Naphthalene)		Thick	300 - 6000	1	20 sec	2	-	Gasoline-30	D5085898	10
Gasoline (Naphthalene)		Thin	300 - 6000	1	2,5 min	1	-	SGA-6L	UNI-CF000636	10
Glycerate 3- Phosphate (GP)	C3H7O7P	Thick	0,0005 - 10 mg/m3	50	10 sec	1	-	CWA- GB.GD.GP.GA.G F.VX-Tubes	7Sol-008	10
Glycerate 3- Phosphate (GP)	C3H7O7P	Thick	0,005 - 100 mg/m3	5	10 sec	1	-	CWA- GB.GD.GP.GA.G F.VX-Tubes	7Sol-008	10
Glycerate 3- Phosphate (GP)	C3H7O7P	Thick	>0,01 mg/m3	10	10 sec	1	-	Phosph. Acid Esters-CWA	D5086880	10
Heptane, All Isomers		Thick	Warning Marks	2 to 15	20 sec	2	-	BNZ-Test	D5085807	10
Heptane, n		Thick	Warning Marks	2	20 sec	2	-	Gasoline-A	D5085769	10
Heptane, n		Thick	50 - 2600	2	20 sec	2	-	Hexane-20	D5086832	10
Heptane, n		Thick	20 - 1000	5	20 sec	2	-	Hexane-20	D5086832	10
Heptane, n		Thin	15 - 1800	1	1,5 min	1	-	SHE-5L	UNI-CF000068	10
Heptane, n		Thin	0,05875 - 1,41 %	1	1,5 min	1	-	SHE-6L	UNI-CF000069	10
Hexane, All Isomers		Thick	Warning Marks	2 to 15	20 sec	2	-	BNZ-Test	D5085807	10
Hexane, n		Thick	Warning Marks	2	20 sec	2	-	Gasoline-A	D5085769	10
Hexane, n		Thick	20 - 1000	5	20 sec	2	-	Hexane-20	D5086832	10
Hexane, n		Thick	80 - 3200	2	20 sec	2	-	Hexane-20	D5086832	10
Hexane, n		Thick	n/a	n/a	n/a	2	-	Hexane-20- ETW	D5093832	10

Hexane, n		Thin	10 - 1200	1	1,5 min	1	-	SHE-5L	UNI-CF000068	10
Hexane, n		Thin	250 - 6000	1	1 min	2	-	SHE-6L	UNI-CF000069	10
Hydrazine		Thick	0,05 - 1,2	10	50 sec	2	-	Methyl Bromide-200	D5086847	10
Hydrazine		Thick	0,1 - 3	5	50 sec	2	-	Hydrazine-0,1	D5093861	10
Hydrazine		Thick	0,5 - 10	2	50 sec	2	-	Hydrazine-0,1	D5093861	10
Hydrazine		Thick	n/a	n/a	n/a	2	-	Hydrazine-0,1-ETW	D5093861	10
Hydrazine		Thin	0.1 - 2	5	1 min	2	H	SHY-1L	UNI-CF000070	10
Hydro Carbons		Thick	20 - 1000	5	20 sec	2	-	Hexane-20	D5086832	10
Hydro Carbons		Thin	10 - 1200	1	1,5 min	1	-	SHE-5L	UNI-CF000068	10
Hydro Carbons		Thin	250 - 6000	1	1 min	2	-	SHE-6L	UNI-CF000069	10
Hydro Carbons (Aromatic)		Thick	5 - 300	10	10 sec	2	-	Aromatic-HC	D5086811	10
Hydro Carbons (Halogenated)		Thick	Warning Marks	n/a	n/a	2	-	CKW-A	D5085771	10
Hydrocyanic Acid	HCL	Thick	n/a	n/a	n/a	n/a	n/a	Hydrocyanic Acid-2-ETW	D5092824	10
Hydrogen		Thin	500 - 8000	½	1 min	1	-	SHD-6M	UNI-CF000659	5
Hydrogen Bromide (Hydrobromic Acid)		Thin	0,8 - 16	1	30 sec	1	-	15L	15L	10
Hydrogen Chloride (Hydrochloric Acid)	HCL	Thick	Warning Marks	2	20 sec	2	-	HCL-A	D5085762	10
Hydrogen Chloride (Hydrochloric Acid)	HCL	Thick	5 - 50	2	20 sec	2	-	HCL-1	D5085846	10
Hydrogen Chloride (Hydrochloric Acid)	HCL	Thick	1 - 10	10	20 sec	2	-	HCL-1	D5085846	10
Hydrogen Chloride (Hydrochloric Acid)	HCL	Thick	5 - 50	2	20 sec	2	-	HCL-1	D5085946	200
Hydrogen Chloride (Hydrochloric Acid)	HCL	Thick	1 - 10	10	20 sec	2	-	HCL-1	D5085946	200
Hydrogen Chloride (Hydrochloric Acid)	HCL	Thick	50 - 500	8	20 sec	2	-	HCL-50	D5086858	10
Hydrogen Chloride (Hydrochloric Acid)	HCL	Thick	500 - 5000	1	20 sec	2	-	HCL-50	D5086858	10
Hydrogen Chloride (Hydrochloric Acid)	HCL	Thick	n/a	n/a	n/a	2	-	HCL-1-ETW	D5092846	10

Hydrogen Chloride (Hydrochloric Acid)	HCL	Thin	1 - 10	3	1 min	2	-	SHC-1	UNI-CF000071	10
Hydrogen Chloride (Hydrochloric Acid)	HCL	Thin	2 - 20	2	2 min	2	-	SHC-2M	UNI-CF000072	5
Hydrogen Chloride (Hydrochloric Acid)	HCL	Thin	5 - 100	1	1 min	1	-	SHC-3	UNI-CF000073	5
Hydrogen Chloride (Hydrochloric Acid)	HCL	Thin	50 - 1000	1	1 min	2	-	SHC-4	UNI-CF000074	10
Hydrogen Cyanide	HCN	Thin	0,3 - 8	3	1,5 min	1	-	112SC	112SC	10
Hydrogen Cyanide	HCN	Thick	Warning Marks	2	20 sec	2	-	HCN-A	D5085756	10
Hydrogen Cyanide	HCN	Thick	5 - 50	2	20 sec	2	-	HCN-2	D5085824	10
Hydrogen Cyanide	HCN	Thick	2 - 12	10	20 sec	2	-	HCN-2	D5085824	10
Hydrogen Cyanide	HCN	Thick	5 - 50	2	20 sec	2	-	HCN-2	D5085924	200
Hydrogen Cyanide	HCN	Thick	2 - 12	10	20 sec	2	-	HCN-2	D5085924	200
Hydrogen Cyanide	HCN	Thin	1,25 - 25	4	1,5 min	2	-	SHN-2	UNI-CF000075	10
Hydrogen Cyanide	HCN	Thin	2,5 - 50	2	1,5 min	2	-	SHN-2	UNI-CF000075	10
Hydrogen Cyanide	HCN	Thin	5 - 100	1	1,5 min	2	-	SHN-2	UNI-CF000075	10
Hydrogen Cyanide	HCN	Thin	2.5 - 50	2	1,5 min	2	-	SHN-3	UNI-CF000076	10
Hydrogen Cyanide	HCN	Thin	5 - 100	1	1,5 min	2	-	SHN-3	UNI-CF000076	10
Hydrogen Cyanide	HCN	Thin	10 - 200	½	1,5 min	2	-	SHN-3	UNI-CF000077	10
Hydrogen Cyanide	HCN	Thin	50 - 1000	1	1 min	2	-	SHN-4	UNI-CF000077	10
Hydrogen Cyanide (AC)	HCN	Thick	1 - 500 mg/m3	15	10 sec	2	-	CWA- CK.AC.DP.CG- Tubes	7Sol-004	10
Hydrogen Cyanide (AC)	HCN	Thick	5 mg/m3	30	10 sec	2	-	CK, AC, DP, CG- CWA	D5086884	10
Hydrogen Fluoride	HF	Thick	5 - 50	2	40 sec	2	-	HF-1	D5086830	10
Hydrogen Fluoride	HF	Thick	1 - 12	8	40 sec	2	-	HF-1	D5086830	10
Hydrogen Fluoride	HF	Thick	n/a	n/a	n/a	2	-	HF-1-ETW	D5093830	10
Hydrogen Fluoride	HF	Thin	0.5 - 30	4	1 min	1	H	SHF-2	UNI-CF000078	10
Hydrogen Peroxide	H2O2	Thick	0,1 - 3	20	10 sec	3	-	Hydrogen Peroxide 0,1/a	8101041	10
Hydrogen Peroxide	H2O2	Thin	0,5 - 10	5	1,5 min	3	-	247S	247S	10
Hydrogen Sulfide	H2S	Thick	0,1 - 6	10	20	2	-	H2S-0,1	10023818	10

Hydrogen Sulfide	H2S	Thick	Warning Marks	2	20 sec	2	-	H2S-A	D5085764	10
Hydrogen Sulfide	H2S	Thick	1 - 200	1	20 sec	3	-	H2S-1	D5085826	10
Hydrogen Sulfide	H2S	Thick	100 - 4000	1	20 sec	3	-	H2S-100	D5085827	10
Hydrogen Sulfide	H2S	Thick	1 - 200	1	20 sec	3	-	H2S-1	D5085926	200
Hydrogen Sulfide	H2S	Thick	100 - 4000	1	20 sec	3	-	H2S-100	D5085927	200
Hydrogen Sulfide	H2S	Thick	0,4 - 4 %	1	20	2	-	H2S-0,4%	D5086857	10
Hydrogen Sulfide	H2S	Thick	n/a	n/a	n/a	2	-	H2S-1-ETW	D5092826	10
Hydrogen Sulfide	H2S	Thin	0.5 - 10	5	1,5 min	3	-	SHS-1	UNI-CF000079	10
Hydrogen Sulfide	H2S	Thin	0.5 - 15	4	1,5 min	3	-	SHS-2	UNI-CF000080	10
Hydrogen Sulfide	H2S	Thin	1 - 30	2	1,5 min	3	-	SHS-2	UNI-CF000080	10
Hydrogen Sulfide	H2S	Thin	2- 60	1	1,5 min	3	-	SHS-2	UNI-CF000080	10
Hydrogen Sulfide	H2S	Thin	1.25 - 30	2	1,5 min	3	-	SHS-3L	UNI-CF000081	10
Hydrogen Sulfide	H2S	Thin	2.5 - 60	1	1,5 min	3	-	SHS-3L	UNI-CF000081	10
Hydrogen Sulfide	H2S	Thin	5 - 120	½	1,5 min	3	-	SHS-3L	UNI-CF000081	10
Hydrogen Sulfide	H2S	Thin	2.5 - 50	2	1,5 min	3	-	SHS-3	UNI-CF000082	10
Hydrogen Sulfide	H2S	Thin	5 - 100	1	1,5 min	3	-	SHS-3	UNI-CF000082	10
Hydrogen Sulfide	H2S	Thin	10 - 200	½	1,5 min	3	-	SHS-3	UNI-CF000082	10
Hydrogen Sulfide	H2S	Thin	25 - 400	2	2,5 min	3	-	SHS-4	UNI-CF000085	10
Hydrogen Sulfide	H2S	Thin	50 - 800	1	2,5 min	3	-	SHS-4	UNI-CF000085	10
Hydrogen Sulfide	H2S	Thin	100 - 1600	½	2,5 min	3	-	SHS-4	UNI-CF000085	10
Hydrogen Sulfide	H2S	Thin	50 - 1000	2	1,5 min	3	-	SHS-5M	UNI-CF000086	10
Hydrogen Sulfide	H2S	Thin	100 - 2000	1	1,5 min	3	-	SHS-5M	UNI-CF000086	10
Hydrogen Sulfide	H2S	Thin	200 - 4000	½	1,5 min	3	-	SHS-5M	UNI-CF000086	10
Hydrogen Sulfide	H2S	Thin	0.05 - 1%	2	2,5 min	3	-	SHS-7M	UNI-CF000087	10
Hydrogen Sulfide	H2S	Thin	0.1 - 2%	1	2,5 min	3	-	SHS-7M	UNI-CF000087	10
Hydrogen Sulfide	H2S	Thin	0.2 - 4%	½	2,5 min	3	-	SHS-7M	UNI-CF000087	10
Hydrogen Sulfide	H2S	Thin	0.25 - 2%	2	1,5 min	3	-	SHS-7	UNI-CF000088	10
Hydrogen Sulfide	H2S	Thin	0.5 - 4%	1	1,5 min	3	-	SHS-7	UNI-CF000088	10
Hydrogen Sulfide	H2S	Thin	1 - 8%	1	1,5 min	3	-	SHS-7	UNI-CF000088	10
Hydrogen Sulfide	H2S	Thin	0.1 - 3.5%	2	1,5 min	3	-	SHS-8L	UNI-CF000089	10
Hydrogen Sulfide	H2S	Thin	0.2 - 7%	1	1,5 min	3	-	SHS-8L	UNI-CF000089	10
Hydrogen Sulfide	H2S	Thin	0.4 - 14%	½	1,5 min	3	-	SHS-8L	UNI-CF000089	10

Hydrogen Sulfide	H2S	Thin	0.25 - 5%	2	2,5 min	3	-	SHS-8M	UNI-CF000090	10
Hydrogen Sulfide	H2S	Thin	0.5 - 10%	1	2,5 min	3	-	SHS-8M	UNI-CF000090	10
Hydrogen Sulfide	H2S	Thin	1 - 20%	1	2,5 min	3	-	SHS-8M	UNI-CF000090	10
Hydrogen Sulfide	H2S	Thin	1 - 20%	1	6 min	3	-	SHS-9	UNI-CF000091	10
Hydrogen Sulfide	H2S	Thin	2 - 40%	½	6 min	3	-	SHS-9	UNI-CF000091	10
Hydrogen Sulfide	H2S	Thin	5 - 125	2	1 min	3	-	SHS-4L	UNI-CF000647	10
Hydrogen Sulfide	H2S	Thin	10 - 250	1	1 min	3	-	SHS-4L	UNI-CF000647	10
Hydrogen Sulfide	H2S	Thin	20 - 500	½	1 min	3	-	SHS-4L	UNI-CF000647	10
Hydrogen Sulfide	H2S	Thin	0.25 - 3.5	2	2 min	1	H	SHS-1H	UNI-CF000727	10
Hydrogen Sulfide	H2S	Thin	0.5 - 7	1	2 min	1	H	SHS-1H	UNI-CF000727	10
Hydrogen Sulfide	H2S	Thin	1 - 14	½	2 min	1	H	SHS-1H	UNI-CF000727	10
Hydrogen Sulfide + Sulfur Dioxide	H2S + SO2	Thin	2.5-60 + 0.5-10	1	2 min	1	-	SHS-3L+SSD-1	UNI-CF000092	5
Hydrogen Sulfide in CNG	H2S	Thin	0.25 - 3.5	2	2 min	1	H	SHS-(1H)	UNI-CF000758	10
Hydrogen Sulfide in CNG	H2S	Thin	0.5 - 7	1	2 min	1	H	SHS-(1H)	UNI-CF000758	10
Hydrogen Sulfide in CNG	H2S	Thin	1 - 14	½	2 min	1	H	SHS-(1H)	UNI-CF000758	10
Iodine		Thin	1.8 - 36	2	1,5 min	1	-	SAG-2	UNI-CF000680	10
Isobutane		Thick	200 - 4200	3	20 sec	2	-	Propane-200	D5086831	10
Isobutyl Amine		Thick	3 - 36	2	20 sec	2	-	Triethyl Amine-5	D5086816	10
Isobutyl Methyl Ketone		Thick	50 - 6500	3	40 sec	2	-	MEK-50	D5086837	10
Isobutylene (Olefins)		Thick	400 - 2600	10	20 sec	2	-	Ethylene-50	D5086833	10
Isooctane		Thick	100 - 3000	?	20 sec	2	-	Hexane-20	D5086832	10
Isopropyl Alcohol (IPA)		Thin	0.1 - 2.5 %	3	1,5 min	2	-	SIP-7M	UNI-CF000093	10
Isopropyl Amine		Thick	3 - 30	2	20 sec	2	-	Triethyl Amine-5	D5086816	10
Isopropyl Amine		Thin	2.5 - 50	1	1,5 min	1	-	SMA-3	UNI-CF000668	10
Isopropyl Ether		Thin	0.013 - 0.34%	3	2 min	2	-	SDE-6	UNI-CF000056	10

Isopropyl Mercaptan		Thick	0,5 - 5,5	15	30 sec	2	-	Ethyl Mercaptan-0,5	D5086815	10
Kerosine		Thick	Warning Marks	-	20 sec	2	-	BNZ-Test	D5085807	10
Kerosine		Thick	Qualitative	1 to 10	10 sec	3	-	QL	D5085910	200
Kerosine (PID detector tube)		Thick	Qualitative	1 to 10	10 sec	3	-	QL	D5085810	10
Ketones		Thick	Qualitative	1 to 10	10 sec	3	-	QL	D5085910	200
Ketones (PID detector tube)		Thick	Qualitative	1 to 10	10 sec	3	-	QL	D5085810	10
Lewisite (L)	C2H2AsCl3	Thick	2 - 1300 mg/m3	40	10 sec	2	-	CWA-Lewite-Tubes	7Sol-005	10
Lewisite (L)	C2H2AsCl3	Thick	1 mg/m3	30	10 sec	2	-	Lewisite-CWA	D5086883	10
Lewisite (L)	C2H2AsCl3	Thick	>1mg/m3	30	10 sec	2	-	CWA-SA.L.DA.DC-Tubes	7Sol-010	10
Light Fuels		Thick	Warning Marks	2 to 15	15 sec	2	-	BNZ-Test	D5085807	10
Ligron		Thick	Warning Marks	2 to 15	20 sec	2	-	BNZ-Test	D5085807	10
Ligron (PID detector tube)		Thick	Qualitative	1 to 10	10 sec	3	-	QL	D5085810	10
Ligron (PID detector tube)		Thick	Qualitative	1 to 10	10 sec	3	-	QL	D5085910	200
Ligroin		Thick	30 - 600	10	20 sec	2	-	Gasoline-30	D5085898	10
Ligroin		Thick	300 - 6000	1	20 sec	2	-	Gasoline-30	D5085898	10
Liquid Petroleum Gas		Thick	200 - 4000	5	20 sec	2	-	Propane-200	D5086831	10
Liquid Petroleum Gas (PID detector tube)		Thick	Qualitative	1 to 10	10 sec	3	-	QL	D5085810	10
Liquid Petroleum Gas (PID detector tube)		Thick	Qualitative	1 to 10	10 sec	3	-	QL	D5085910	200
Lost-S (Yperite, HD)	C4H8Cl2S	Thick	>1mg/m3	50	10 sec	2	-	CWA-HD-Tubes	7Sol-003	10

Lost-S (Yperite, HD)	C4H8Cl2S	Thick	n/a	n/a	n/a	2	-	HD-CWA	D5086882	10
Lost-S (Yperite, Nitrogen Mustard, H, HN, HD)		Thick	1 mg/m3	30	10 sec	2	-	HD, HN-CWA	D5086881	10
Maleic Anhydride		Thin	1 - 20	1	1 min	1	-	SAA-2	UNI-CF000001	10
Mercaptans		Thick	0,1 - 1	75	30 sec	2	-	Ethyl Mercaptan-0,5	D5086815	10
Mercaptans		Thick	0,5 - 5	15	30 sec	2	-	Ethyl Mercaptan-0,5	D5086815	10
Mercaptans		Thick	10 - 100	1	30 sec	2	-	Ethyl Mercaptan-0,5	D5086815	10
Mercaptans		Thin	0.5 - 10	2	2 min	1	-	SEM-2M	UNI-CF000059	10
Mercaptans		Thin	1 - 20	1	3 min	1	-	SEM-2M	UNI-CF000059	10
Mercaptans		Thin	2 - 40	½	3 min	1	-	SEM-2M	UNI-CF000059	10
Mercaptans		Thin	2.5 - 60	2	1,5 min	1	-	SEM-4L	UNI-CF000060	10
Mercaptans		Thin	5 - 120	1	1,5 min	1	-	SEM-4L	UNI-CF000060	10
Mercaptans		Thin	10 - 240	½	1,5 min	1	-	SEM-4L	UNI-CF000060	10
Mercaptans		Thin	12.5 - 250	2	1,5 min	1	-	SEM-4M	UNI-CF000061	10
Mercaptans		Thin	25 - 500	1	1,5 min	1	-	SEM-4M	UNI-CF000061	10
Mercaptans		Thin	0.5 - 30 mg/m3 (0.1 - 8 ppm)	2	1 min	1	-	SBM-1	UNI-CF000138	10
Mercaptans		Thin	0.5 - 15 mg/m3 (0.1 - 4 ppm)	3	1,5 min	1	-	SBM-1M	UNI-CF000694	10
Mercaptans		Thin	2 - 80	2	1 min	1	-	SEM-(3L)	UNI-CF000784	10
Mercury Vapour	Hg	Thick	0,048 - 0,384 (0,4 - 3,2 mg/m3)	5	25 sec	2	-	Hg-0,01	D5085843	10
Mercury Vapour	Hg	Thick	0,012 - 0,096 (0,1 - 0,8 mg/m3)	20	25 sec	2	-	Hg-0,01	D5085843	10

Mercury Vapour	Hg	Thick	0,003 - 0,024 (0,025 - 0,2 mg/m3)	80	25 sec	2	-	Hg-0,01	D5085843	10
Mercury Vapour	Hg	Thin	0.1 - 2 mg/m3	1	2 min	1	-	SHG-1L	UNI-CF000097	10
Methane (Natural Gas)	CH4	Thick	From 0,5% Warning Marks	5	2 min	1	-	Natural Gas	D5086846	10
Methanol (Methyl Alcohol)		Thick	100 - 2350	10	15 sec	2	-	Ethanol-100	D5086818	10
Methanol (Methyl Alcohol)		Thick	n/a	n/a	n/a	2	-	Ethanol-100- ETW	D5094818	10
Methyl Acetate		Thin	0,1 - 3 %	1	1,5 min	3	-	11SA	11SA	10
Methyl Acrylate		Thick	5 - 200	20	15 sec	3	-	Methyl Acrylate 5/a	6728161	10
Methyl Alcohol (Methanol)		Thick	Warning Marks	15 sec	20 sec	2	-	Alcohols-A	D5085773	10
Methyl Alcohol (Methanol)		Thick	100 - 2500	n/a	n/a	2	-	Methyl Alcohol- 100	D5094818	10
Methyl Alcohol (Methanol)		Thin	100 - 5000	1	1,5 min	2	-	SME-5	UNI-CF000098	10
Methyl Alcohol (Methanol)		Thin	0.05 - 6 %	2	2 min	2	-	SME-8L	UNI-CF000099	10
Methyl Amine		Thick	4 - 55	1	20 sec	2	-	Triethyl Amine- 5	D5086816	10
Methyl Amine		Thin	5 - 100	1	1 min	1	-	SMA-3	UNI-CF000668	10
Methyl Bromide (Dibromethane)		Thick	0,2 - 2	5	1,5 min	3	-	Methyl Bromide 0,2/a	8103391	10
Methyl Bromide (Dibromethane)		Thick	0,2 - 8	2	1,5 min	3	-	Methyl Bromide 0,2/a	8103391	10
Methyl Bromide (Dibromethane)		Thick	9 - 200	10	20 sec	2	-	Trichloroethan e-5	D5086834	10
Methyl Bromide (Dibromethane)		Thick	2 - 10	6	n/a	2	-	Methyl Bromide -2	D5086845	5
Methyl Bromide (Dibromethane)		Thick	10 - 100	1	n/a	2	-	Methyl Bromide-2	D5086845	5

Methyl Bromide (Dibromethane)		Thick	2000 - 20000	1	20 sec	2	-	Methyl Bromide-200	D5086847	10
Methyl Bromide (Dibromethane)		Thick	200 - 8000	2	20 sec	2	-	Methyl Bromide-200	D5086847	10
Methyl Bromide (Dibromethane)		Thin	0.5 - 1	3	2,5 min	1	-	SMB-2M	UNI-CF000100	5
Methyl Bromide (Dibromethane)		Thin	1 - 18	2	2,5 min	1	-	SMB-2M	UNI-CF000100	5
Methyl Bromide (Dibromethane)		Thin	2 - 36	1	2,5 min	1	-	SMB-2M	UNI-CF000100	5
Methyl Bromide (Dibromethane)		Thin	5 - 80	½	2,5 min	1	-	SMB-2M	UNI-CF000100	5
Methyl Bromide (Dibromethane)		Thin	5 - 50	2	2 min	1	-	SMB-3	UNI-CF000101	5
Methyl Bromide (Dibromethane)		Thin	10 - 100	1	3 min	1	-	SMB-3	UNI-CF000101	5
Methyl Bromide (Dibromethane)		Thin	20 - 200	½	4 min	1	-	SMB-3	UNI-CF000101	5
Methyl Bromide (Dibromethane)		Thin	25 - 500	2	2,5 min	1	-	SMB-4	UNI-CF000103	5
Methyl Bromide (Dibromethane)		Thin	50 - 1000	1	2,5 min	1	-	SMB-4	UNI-CF000103	5
Methyl Bromide (Dibromethane)		Thin	100 - 2000	½	2,5 min	1	-	SMB-4	UNI-CF000103	5
Methyl Bromide (Dibromethane)		Thin	10 - 150	2	1,5 min	1	-	SMB-4M	UNI-CF000402	5
Methyl Bromide (Dibromethane)		Thin	20 - 300	1	1,5 min	1	-	SMB-4M	UNI-CF000402	5
Methyl Bromide (Dibromethane)		Thin	20 - 600	½	1,5 min	1	-	SMB-4M	UNI-CF000402	5
Methyl Butane, 2		Thick	50 - 3000	2	20 sec	2	-	Hexane-20	D5086832	10
Methyl Cycloheptane		Thick	80 - 4900	5	20 sec	2	-	Hexane-20	D5086832	10
Methyl Cyclohexane		Thin	10 - 1600	3	2,5 min	1	-	SMY-5M	UNI-CF000782	10
Methyl Cyclopentane		Thick	150 - 3700	2	20 sec	2	-	Hexane-20	D5086832	10

Methyl Ethyl Ketone (MEK)		Thin	0.02 - 0.6%	2	1,5 min	2	-	SMK-6L	UNI-CF000104	10
Methyl Ethyl Ketone (MEK)		Thick	50 - 4000	3	40 sec	2	-	MEK-50	D5086837	10
Methyl Hydrazine		Thick	0,2 - 4	5	50 sec	2	-	Hydrazine-0,1	D5086861	10
Methyl Iodide		Thin	500 - 15000	1	1 min	1	-	SMO-7M	UNI-CF000779	10
Methyl Isobutyl Ketone (MIBK)		Thin	0.02 - 0.6 %	2	1,5 min	2	-	SMI-6L	UNI-CF000107	10
Methyl Mercaptan		Thick	0,5 - 5	15	30 sec	2	-	Ethyl Mercaptan-0,5	D5086815	10
Methyl Mercaptan		Thin	0.5 - 10	2	1,5 min	1	-	SEM-2M	UNI-CF000059	10
Methyl Mercaptan		Thin	1 - 20	1	1,5 min	1	-	SEM-2M	UNI-CF000059	10
Methyl Mercaptan		Thin	2 - 40	½	1,5 min	1	-	SEM-2M	UNI-CF000059	10
Methyl Mercaptan		Thin	2.5 - 60	2	1,5 min	1	-	SEM-4L	UNI-CF000060	10
Methyl Mercaptan		Thin	5 - 120	1	1,5 min	1	-	SEM-4L	UNI-CF000060	10
Methyl Mercaptan		Thin	10 - 240	½	1,5 min	1	-	SEM-4L	UNI-CF000060	10
Methyl Mercaptan		Thin	12.5 - 250	2	1,5 min	1	-	SEM-4M	UNI-CF000061	10
Methyl Mercaptan		Thin	25 - 500	1	1,5 min	1	-	SEM-4M	UNI-CF000061	10
Methyl Mercaptan		Thin	0.5 - 30 mg/m3 (0.1 8 ppm)	2	1 min	1	-	SBM-1	UNI-CF000138	10
Methyl Mercaptan		Thin	0.5 - 15 mg/m3 (0.1 - 4 ppm)	3	1,5 min	1	-	SBM-1M	UNI-CF000694	10
Methyl Mercaptan		Thin	2 - 80	1	1,5 min	1	-	SEM-(3L)	UNI-CF000784	10
Methyl Methacrylate		Thick	1,5 - 115	3	20 sec	2	-	Acrolein-0,1	10017882	10
Methyl Pentane, 2		Thick	150 - 4500	2	20 sec	2	-	Hexane-20	D5086832	10
Methyl Pentane, 3		Thick	100 - 3700	2	20 sec	2	-	Hexane-20	D5086832	10
Methylene Chloride		Thin	50 - 500	1	2,5 min	1	-	SMC-4M	UNI-CF000639	5
Methylene Chloride		Thin	30 - 1000	2	2 min	1	-	SMC-4	UNI-CF000762	5

Methylene Chloride		Thin	10 - 60	2	2,5 min	1	-	SMC-3L	UNI-CF000793	5
Mineralic Oil / Synthetic Oil (Compressor Oil Mist)		Thick	Warning Marks	-	-	2	-	Oil-HP	10040887 (before D5085850)	10
Morpholine		Thin	10 - 200	1	1 min	1	-	SMA-3	UNI-CF000668	10
Multigas Detector: Ammonia Amines Sulfur Dioxide Acetic Acid Hydrogen Chloride Chlorine Nitrogen Dioxide Hydrogen Sulphide Carbon Monoxide Phosphine Acetylene Mercaptans	-	Thin	See measuring ranges at UNIPHOS.eu	1	1,5 min	1	-	SMG-1	UNI-CF000697	10
Natural Gas (Methane)	CH4	Thick	From 0,5% Warning Marks	5	2 min	1	-	Natural Gas	D5086846	10
Nickel Tetracarbonyl	Ni(CO)4	Thick	0,1 - 1	20	15 sec	3	-	Nickel Tetracarbonyl 0,1/a	CH19501	9
Nitric Acid	HNO3	Thin	1 - 20	1	1 min	1	H	SNA-2M	UNI-CF000110	10
Nitrogen Dioxide	NO2	Thick	0,5 - 50	5	20 sec	2	-	NO2-0,5	D5085805	10
Nitrogen Dioxide	NO2	Thick	2 - 140	n/a	n/a	2	-	NO2-2	D5085867	10
Nitrogen Dioxide	NO2	Thick	2 - 140	n/a	n/a	2	-	NO2-2	D5085967	200
Nitrogen Dioxide	NO2	Thick	n/a	n/a	n/a	2	-	NO2-0,5-ETW	D5092805	10
Nitrogen Dioxide	NO2	Thin	0.5 - 30	2	2 min	1	-	SND-2	UNI-CF000111	10
Nitrogen Dioxide	NO2	Thin	10 - 1000	1	1,5 min	1	-	SND-4	UNI-CF000112	10
Nitrogen Oxides (Nitrous Gases/Fumes)	NOx	Thick	Warning Marks	2	20 sec	2	-	Nitr.-A	D5085760	10

Nitrogen Oxides (Nitrous Gases/Fumes)	NOx	Thick	10 - 300	2	20 sec	2	-	Nitr.-10	D5085808	10
Nitrogen Oxides (Nitrous Gases/Fumes)	NOx	Thick	50 - 3000	1	20 sec	2	-	Nitr.-50	D5085809	10
Nitrogen Oxides (Nitrous Gases/Fumes)	NOx	Thick	0,5 - 50	5	20 sec	2	-	Nitr.-0,5	D5085818	10
Nitrogen Oxides (Nitrous Gases/Fumes)	NOx	Thick	5 - 140	5	20 sec	2	-	Nitr.-2	D5085844	10
Nitrogen Oxides (Nitrous Gases/Fumes)	NOx	Thick	2 - 70	10	20 sec	2	-	Nitr.-2	D5085844	10
Nitrogen Oxides (Nitrous Gases/Fumes)	NOx	Thick	50 - 3000	1	20 sec	2	-	Nitr.-50	D5085909	200
Nitrogen Oxides (Nitrous Gases/Fumes)	NOx	Thick	0,5 - 50	5	20 sec	2	-	Nitr.-0,5	D5085918	200
Nitrogen Oxides (Nitrous Gases/Fumes)	NOx	Thick	0,2 - 2,5	3 liter	10 min	2	-	Nitr.-HP	D5086850	10
Nitrogen Oxides (Nitrous Gases/Fumes)	NOx	Thick	0,5 - 6	1,5 liter	5 min	2	-	Nitr.-HP	D5086850	10
Nitrogen Oxides (Nitrous Gases/Fumes)	NOx	Thin	0.5 - 15	2	1,5 min	1	-	SNO-2M	UNI-CF000113	10
Nitrogen Oxides (Nitrous Gases/Fumes)	NOx	Thin	100 - 2500	1	1,5 min	1	-	SNO-5M	UNI-CF000114	10
Nitrogen Oxides (Nitrous Gases/Fumes)	NOx	Thin	20 - 250	1	1,5 min	1	-	SNO-4L	UNI-CF000662	10
Nonane, n		Thick	Warning Marks	2	20 sec	2	-	Gasoline-A	D5085769	10
Nonane, n		Thick	50 - 2800	10	20 sec	2	-	Hexane-20	D5086832	10

Octane, n	CH3(CH2)6 CH3	Thick	Warning Marks	2	20 sec	2	-	Gasoline-A	D5085769	10
Octane, n	CH3(CH2)6 CH3	Thick	50 - 3000	5	20 sec	2	-	Hexane-20	D5086832	10
Octane, n	CH3(CH2)6 CH3	Thin	23 - 2760	1	1,5 min	1	-	SHE-5L	UNI-CF000068	10
Octane, n	CH3(CH2)6 CH3	Thin	0,095 - 2,28	1	1 min	1	-	SHE-6L	UNI-CF000069	10
Organic Compounds (PID detector tube)	-	Thick	Qualitative	1 to 10	10 sec	3	-	QL	D5085810	10
Organic Compounds (PID detector tube)	-	Thick	Qualitative	1 to 10	10 sec	3	-	QL	D5085910	200
Otto Fuel		Thick	n/a	n/a	n/a	30 Mon.	n/a	Otto Fuel	D5086817	n/a
Oxygen	O2	Thin	3 - 24 %	½	4 min	2	-	SOX-9M	UNI-CF000116	5
Ozone	O3	Thick	0,05 - 5	2	40 sec	2	-	Ozone-0,05	D5086828	10
Ozone	O3	Thin	5 - 100	1	1,5 min	2	-	SOZ-3	UNI-CF000117	10
Ozone	O3	Thin	25 - 500	1	1,5 min	2	-	SOZ-2M	UNI-CF000118	10
Pentachloroethane		Thick	10 - 300	10	20 sec	2	-	Trichloroethan e-5	D5086834	10
Pentane, n	C5H12	Thick	Warning Marks	2	20 sec	2	-	Gasoline-A	D5085769	10
Pentane, n	C5H12	Thick	350 - 3900	1	20 sec	2	-	Hexane-20	D5086832	10
Pentane, n	C5H12	Thick	50 - 500	5	20 sec	2	-	Hexane-20	D5086832	10
Pentane, n	C5H12	Thin	7 - 840	1	1,5 min	1	-	SHE-5L	UNI-CF000068	10
Pentane, n	C5H12	Thin	85 - 2040	1	1 min	2	-	SHE-6L	UNI-CF000069	10
Petroleum Distillates (PID detector tube)	-	Thick	Qualitative	1 to 10	10 sec	3	-	QL	D5085810	10
Petroleum Distillates (PID detector tube)	-	Thick	Qualitative	1 to 10	10 sec	3	-	QL	D5085910	200
Petroleum Distillates (Naphthalene)	-	Thick	Warning Marks	2 to 15	20 sec	2	-	BNZ-Test	D5085807	10

Petroleum Distillates (Naphthalene)	-	Thick	300 - 6000	1	20 sec	2	-	Gasoline-30	D5085898	10
Petroleum Distillates (Naphthalene)	-	Thick	30 - 600	10	20 sec	2	-	Gasoline-30	D5085898	10
Petroleum Ether		Thick	Warning Marks	2 to 15	20 sec	2	-	BNZ-Test	D5085807	10
Phenol	C6H6O	Thick	1 - 15	20	15 sec	2	-	Phenol-1	D5086838	10
Phenol	C6H6O	Thick	5 - 25	10	15 sec	2	-	Phenol-1	D5086838	10
Phenol	C6H6O	Thick				2	-	Phenol-1-ETW	D5093838	10
Phenol	C6H6O	Thin	1 - 25	2	2 min	1	-	SPE-2M	UNI-CF000119	10
Phosgene (Carbonyl Dichloride)	CCl2O	Thick	Warning Marks	5	20 sec	2	-	Phosgene-A	D5085761	10
Phosgene (Carbonyl Dichloride)	CCl2O	Thick	1 - 20	5	20 sec	2	-	Phosgene-0,1	D5085854	10
Phosgene (Carbonyl Dichloride)	CCl2O	Thick	0,1 - 5	20	20 sec	2	-	Phosgene-0,1	D5085854	10
Phosgene (Carbonyl Dichloride)	CCl2O	Thick	n/a	n/a	n/a	2	-	Phosgene-0,1- ETW	D5092854	10
Phosgene (Carbonyl Dichloride)	CCl2O	Thin	0.1 - 10	1	2 min	1	-	SPG-1	UNI-CF000120	10
Phosgene (CG)	COCl2	Thick	1 - 2000 mg/m3	15	10 sec	2	-	CWA- CK.AC.DP.CG- Tubes	7Sol-004	10
Phosgene (CG)	COCl2	Thick	5 mg/m3	30	10 sec	2	-	CK, AC, DP, CG- CWA	D5086884	10
Phosphine	PH3	Thick	n/a	n/a	n/a	n/a	n/a	PH3-0,01	10098412	n/a
Phosphine	PH3	Thick	0,1 - 3,0	10	25 sec	2	-	PH3-0,05	D5085829	10
Phosphine	PH3	Thick	0,05 - 1,5	20	25 sec	2	-	PH3-0,05	D5085829	10
Phosphine	PH3	Thick	0,1 - 10	10	20 sec	3	-	PH3-0,1	D5085830	10
Phosphine	PH3	Thick	50 - 2000	1	20 sec	3	-	PH3-50	D5085831	10

Phosphine	PH3	Thick	0,05 - 1,5	20	25 sec	2	-	PH3-0,05	D5085929	200
Phosphine	PH3	Thick	0,05 - 1,5	20	25 sec	2	-	PH3-0,05	D5085929	200
Phosphine	PH3	Thick	0,1 - 10	10	20 sec	3	-	PH3-0,1	D5085930	200
Phosphine	PH3	Thick	50 - 2000	1	20 sec	3	-	PH3-50	D5085931	200
Phosphine	PH3	Thick	n/a	n/a	n/a	2	-	PH3-0,05-ETW	D5092829	10
Phosphine	PH3	Thin	0.05 - 2.5	2	2 min	2	-	SPH-1L	UNI-CF000123	10
Phosphine	PH3	Thin	0.1 - 5	1	2 min	2	-	SPH-1L	UNI-CF000123	10
Phosphine	PH3	Thin	0.05 - 5	2	1,5 min	2	-	SPH-1	UNI-CF000125	10
Phosphine	PH3	Thin	0.1 -10	1	2 min	2	-	SPH-1	UNI-CF000125	10
Phosphine	PH3	Thin	0.2 - 20	½	2 min	2	-	SPH-1	UNI-CF000125	10
Phosphine	PH3	Thin	2.5 - 50	2	1,5 min	2	-	SPH-3	UNI-CF000126	10
Phosphine	PH3	Thin	5 -100	1	1,5 min	2	-	SPH-3	UNI-CF000126	10
Phosphine	PH3	Thin	10 - 200	½	1,5 min	2	-	SPH-3	UNI-CF000126	10
Phosphine	PH3	Thin	25 - 500	2	1,5 min	2	-	SPH-4	UNI-CF000127	10
Phosphine	PH3	Thin	50 - 1000	1	1,5 min	2	-	SPH-4	UNI-CF000127	10
Phosphine	PH3	Thin	100 - 2000	½	1,5 min	2	-	SPH-4	UNI-CF000127	10
Phosphine	PH3	Thin	75 - 1500	1	2 min	2	-	SPH-5	UNI-CF000128	10
Phosphine	PH3	Thin	150 -3000	½	2 min	2	-	SPH-5	UNI-CF000128	10
Phosphine	PH3	Thin	150 - 4000	1	2 min	2	-	SPH-6M	UNI-CF000738	10
Phosphine	PH3	Thin	300 -8000	½	2 min	2	-	SPH-6M	UNI-CF000738	10
Propane	C3H8	Thick	200 - 4000	5	20 sec	2	-	Propane-200	D5086831	10
Propyl Alcohol, iso (Isopropanol)		Thick	Warning Marks	15 sec	20 sec	2	-	Alcohols-A	D5085773	10
Propyl Alcohol, iso (Isopropanol)		Thick	200 - 5000	10	15 sec	2	-	Ethanol-100	D5086818	10
Propyl Alcohol, n		Thick	Warning Marks	15 sec	20 sec	2	-	Alcohols-A	D5085773	10
Propyl Alcohol, n		Thick	100 - 3000	10	15 sec	2	-	Ethanol-100	D5086818	10
Propyl Amine, iso		Thick	5 - 30	2	20 sec	2	-	Triethyl Amine- 5	D5086816	10
Propyl Amine, n		Thick	2 - 28	2	20 sec	2	-	Triethyl Amine- 5	D5086816	10

Propyl Mercaptan, iso		Thick	0,5 - 5,5	15	30 sec	2	-	Ethyl Mercaptan-0,5	D5086815	10
Propyl Mercaptan, n		Thick	0,7 - 8	15	30 sec	2	-	Ethyl Mercaptan-0,5	D5086815	10
Propylene (Olefins)	C3H6	Thick	Warning Marks	2	20 sec	2	-	Gasoline-A	D5085769	10
Propylene (Olefins)	C3H6	Thick	20 - 5000	10	20 sec	2	-	Ethylene-50	D5086833	10
Prussic Acid (see D5093884 or 7Sol-004))	-	Thick	-	-	-	-	-	-	-	-
Pyridine		Thin	1 - 14	1	1,5 min	2	-	SPY-2M	UNI-CF000766	10
Qualitest (PID detector tube)	-	Thick	Qualitative	1 to 10	10 sec	3	-	QL	D5085810	10
Qualitest (PID detector tube)	-	Thick	Qualitative	1 to 10	10 sec	3	-	QL	D5085910	200
Sarin (GB)	C4H10FO2 P	Thick	>0,0005 mg/m3	60	10 sec	5	-	CWA Sarin Tubes	7Sol-002	10
Sarin (GB)	C4H10FO2 P	Thick	>0,0005 mg/m3	50	10 sec	2	-	CWA-GB.GX.VX-Tubes	7Sol-002	10
Sarin (GB)	C4H10FO2 P	Thick	>0,5 mg/m3	5	10 sec	2	-	CWA-GB.GX.VX-Tubes	7Sol-002	10
Sarin (GB)	C4H10FO2 P	Thick	0,0005 - 10 mg/m3	50	10 sec	1	-	CWA-GB.GD.GP.GA.G F.VX-Tubes	7Sol-008	10
Sarin (GB)	C4H10FO2 P	Thick	0,005 - 100 mg/m3	5	10 sec	1	-	CWA-GB.GD.GP.GA.G F.VX-Tubes	7Sol-008	10
Sarin (GB)	C4H10FO2 P	Thick	>0,01 mg/m3	10	10 sec	1	-	Phosph. Acid Esters-CWA	D5086880	10
Scentinel E		Thick	0,8 - 7,5	10	25 sec	2	-	Ethyl Mercaptan-0,5	D5086815	10

Scentinel E		Thick	1,6 - 16,7	5	25 sec	2	-	Ethyl Mercaptan-0,5	D5086815	10
Soman (GD)		Thick	>0,0005 mg/m3	60	10 sec	5	-	CWA Sarin Tubes	7Sol-002	10
Soman (GD)	C7H16FO2 P	Thick	>0,0005 mg/m3	50	10 sec	2	-	CWA-GB.GX.VX-Tubes	7Sol-002	10
Soman (GD)	C7H16FO2 P	Thick	>0,5 mg/m3	5	10 sec	2	-	CWA-GB.GX.VX-Tubes	7Sol-002	10
Soman (GD)	C7H16FO2 P	Thick	0,0005 - 10 mg/m3	50	10 sec	1	-	CWA-GB.GD.GP.GA.G F.VX-Tubes	7Sol-008	10
Soman (GD)	C7H16FO2 P	Thick	0,005 - 100 mg/m3	5	10 sec	1	-	CWA-GB.GD.GP.GA.G F.VX-Tubes	7Sol-008	10
Soman (GD)	C7H16FO2 P	Thick	>0,01 mg/m3	10	10 sec	1	-	Phosph. Acid Esters-CWA	D5086880	10
Stoddard Solvent		Thin	50 - 8000 mg/m3	1	1 min	1	-	SSS-5M	UNI-CF000129	10
Styrene	C8H8	Thin	1 - 50	4	1 min	3	-	158SB	158SB	5
Styrene	C8H8	Thin	2 - 100	2	1 min	3	-	158SB	158SB	5
Styrene	C8H8	Thick	10 - 300	10	10 sec	2	-	Styrene-10	D5086819	10
Styrene	C8H8	Thick	n/a	n/a	n/a	2	-	Styrene-10-ETW	D5093819	10
Styrene	C8H8	Thin	20 - 250	1	1 min	1	-	SST-4L	UNI-CF000787	10
Sulfur Dioxide	SO2	Thick	Warning Marks	n/a	n/a	2	-	SO2-A	D5085763	10
Sulfur Dioxide	SO2	Thick	0,5 - 8	10	25 sec	2	-	SO2-1	D5085803	10
Sulfur Dioxide	SO2	Thick	1 - 25	3	25 sec	2	-	SO2-1	D5085803	10
Sulfur Dioxide	SO2	Thick	40 - 200	1	25 sec	2	-	SO2-5	D5085813	10
Sulfur Dioxide	SO2	Thick	5 - 50	3	25 sec	2	-	SO2-5	D5085813	10
Sulfur Dioxide	SO2	Thick	100 - 500	5	20 sec	2	-	SO2-100	D5085825	10
Sulfur Dioxide	SO2	Thick	500 - 4000	1	20 sec	2	-	SO2-100	D5085825	10
Sulfur Dioxide	SO2	Thick	n/a	n/a	n/a	2	-	SO2-ETW	D5092803	10
Sulfur Dioxide	SO2	Thin	0.25 - 5	4	1,5 min	1	-	SSD-1	UNI-CF000130	10

Sulfur Dioxide	SO2	Thin	0.5 - 10	2	1,5 min	1	-	SSD-1	UNI-CF000130	10
Sulfur Dioxide	SO2	Thin	1 - 20	1	1,5 min	1	-	SSD-1	UNI-CF000130	10
Sulfur Dioxide	SO2	Thin	0.5 - 12.5	4	1,5 min	2	-	SSD-2M	UNI-CF000131	10
Sulfur Dioxide	SO2	Thin	1 - 25	2	1,5 min	2	-	SSD-2M	UNI-CF000131	10
Sulfur Dioxide	SO2	Thin	2 - 50	1	1,5 min	2	-	SSD-2M	UNI-CF000131	10
Sulfur Dioxide	SO2	Thin	2.5 - 50	2	2 min	2	-	SSD-3	UNI-CF000132	10
Sulfur Dioxide	SO2	Thin	5 -100	1	2 min	2	-	SSD-3	UNI-CF000132	10
Sulfur Dioxide	SO2	Thin	10 - 200	½	2 min	2	-	SSD-3	UNI-CF000132	10
Sulfur Dioxide	SO2	Thin	10 - 150	2	1,5 min	2	-	SSD-4M	UNI-CF000133	10
Sulfur Dioxide	SO2	Thin	20 -300	1	1,5 min	2	-	SSD-4M	UNI-CF000133	10
Sulfur Dioxide	SO2	Thin	40 - 600	½	1,5 min	2	-	SSD-4M	UNI-CF000133	10
Sulfur Dioxide	SO2	Thin	250 - 4000	2	1,5 min	2	-	SSD-6M	UNI-CF000134	10
Sulfur Dioxide	SO2	Thin	500 - 8000	1	1,5 min	2	-	SSD-6M	UNI-CF000134	10
Sulfur Dioxide	SO2	Thin	0.1 - 1.6%	½	1,5 min	2	-	SSD-6M	UNI-CF000134	10
Sulfur Dioxide	SO2	Thin	200 - 3000	1	1,5 min	2	-	SSD-5	UNI-CF000135	10
Sulfur Dioxide	SO2	Thin	0.05 - 1.5%	2	1,5 min	2	-	SSD-7M	UNI-CF000136	10
Sulfur Dioxide	SO2	Thin	0.1 - 3 %	1	1,5 min	2	-	SSD-7M	UNI-CF000136	10
Sulfur Dioxide	SO2	Thin	0.2 - 6%	½	1,5 min	2	-	SSD-7M	UNI-CF000136	10
Sulfur Dioxide	SO2	Thin	0.125 - 2,5%	2	1,5 min	2	-	SSD-7	UNI-CF000137	10
Sulfur Dioxide	SO2	Thin	0.25 - 5%	1	1,5 min	2	-	SSD-7	UNI-CF000137	10
Sulfur Dioxide	SO2	Thin	0.5 - 10%	½	1,5 min	2	-	SSD-7	UNI-CF000137	10
Sulfur Hexafluoride	SF6	Thick	0,5 - 15	n/a	n/a	n/a	n/a	SF6	D5085838	10
Sulfuric Acid	H2SO4	Thin	0.5 - 5 mg/m3	5	3 min	1	-	SSA-1L	UNI-CF000700	10
Sulfuryl Fluoride	SO2F2	Thick	1 - 50	6	30 sec	3	-	Sulfuryl Fluoride 1/a	8103471	10
Tabun (GA)	C5H11N2O2P	Thick	0,0005 - 10 mg/m3	50	10 sec	1	-	CWA-GB.GD.GP.GA.G.F.VX-Tubes	7Sol-008	10
Tabun (GA)	C5H11N2O2P	Thick	0,005 - 100 mg/m3	5	10 sec	1	-	CWA-GB.GD.GP.GA.G.F.VX-Tubes	7Sol-008	10
Tabun (GA)	C5H11N2O2P	Thick	>0,01 mg/m3	10	10 sec	1	-	Phosph. Acid Esters-CWA	D5086880	10

Tetrabromoethane	C2H2Br4	Thick	Warning Marks	n/a	n/a	2	-	CKW-A	D5085771	10
Tetrabromoethane	C2H2Br4	Thick	10 - 500	5	30 sec	2	-	Per-5	D5085865	10
Tetrabromoethane	C2H2Br4	Thick	5 - 200	10	30 sec	2	-	Per-5	D5085865	10
Tetrabromoethane, 1,1,2,2	C2H2Br4	Thick	5 - 200	10	20 sec	2	-	Trichloroethane-5	D5086834	5
Tetrachloroethane, 1,1,2,2	C2H2Br4	Thick	50 - 1000	10	20 sec	2	-	Trichloroethane-5	D5086834	5
Tetrachloroethylene (Perchloroethylene)	C2Cl4	Thick	10 - 500	5	30 sec	2	-	Per-10	D5085840	10
Tetrachloroethylene (Perchloroethylene)	C2Cl4	Thick	5 - 200	5	30 sec	2	-	Per-10	D5085840	10
Tetrachloroethylene (Perchloroethylene)	C2Cl4	Thick	10 - 500	5	30 sec	2	-	Per-5	D5085865	10
Tetrachloroethylene (Perchloroethylene)	C2Cl4	Thick	5 - 200	5	30 sec	2	-	Per-5	D5085865	10
Tetrachloroethylene (Perchloroethylene)	C2Cl4	Thick	5 - 200	5	30 sec	2	-	Per-10	D5085940	200
Tetrachloroethylene (Perchloroethylene)	C2Cl4	Thick	n/a	n/a	n/a	2	-	Per-10-ETW	D5092865	10
Tetrachloroethylene (Perchloroethylene)	C2Cl4	Thin	2.5 - 50	2	1,5 min	1	-	STE-3	UNI-CF000685	5
Tetrachloroethylene (Perchloroethylene)	C2Cl4	Thin	5 - 100	1	1,5 min	1	-	STE-3	UNI-CF000685	10
Tetrachloroethylene (Perchloroethylene)	C2Cl4	Thin	10 - 200	½	1,5 min	1	-	STE-3	UNI-CF000685	10
Tetrahydrofuran (THF)	C4H8O	Thick	30 - 980	?	?	2	-	Ethanol-100	D5086818	10
Tetrahydrofuran (THF)	C4H8O	Thin	50 - 800	2	3,5 min	1	-	STH-4	UNI-CF000140	10
Tetrahydrothiophene		Thick	1 - 10	30	30 sec	3	-	Tetrahydrothiophene 1/b	8101341	5
Tetrahydrothiophene		Thick	1 - 10	20	25 sec	2	T	THT-1	D5085866	10
Tetrahydrothiophene		Thick	1 - 10	20	25 sec	2	T	THT-1	D5085966	10
Toluene (Methyl Benzene)	C7H8	Thick	Warning Marks	n/a	n/a	2	-	Aromatic-A	D5085770	10
Toluene (Methyl Benzene)	C7H8	Thick	5 - 1000	5	40 sec	3	-	Tol-5	D5085828	10

Toluene (Methyl Benzene)	C7H8	Thick	30 - 750	10	15 sec	2	-	Aromatic-HC	D5086811	10
Toluene (Methyl Benzene)	C7H8	Thick	n/a	n/a	n/a	3	-	Tol-5-ETW	D5092828	10
Toluene (Methyl Benzene)	C7H8	Thin	1 - 25	4	1 min	1	-	STO-2	UNI-CF000141	10
Toluene (Methyl Benzene)	C7H8	Thin	2 - 50	2	1 min	1	-	STO-2	UNI-CF000141	10
Toluene (Methyl Benzene)	C7H8	Thin	4 - 100	1	1 min	1	-	STO-2	UNI-CF000141	10
Toluene (Methyl Benzene)	C7H8	Thin	10 - 300	1	1,5 min	2	-	STO-4M	UNI-CF000142	10
Toluene Diisocyanate	C9H6N2O2	Thick	0,02 - 0,2	25	50 sec	3		Toluene Diisocyanate 0.02/A	6724501	9
Tri Ethyl Amine		Thin	2.15 - 43	1	1 min	1	-	SMA-3	UNI-CF000668	10
Tri Methyl Amine		Thick	5 - 30	1	20 sec	2	-	Triethyl Amine-5	D5086816	10
Tri Methyl Amine		Thin	1.7 - 43	1	1 min	1	-	SMA-3	UNI-CF000668	10
Tribromomethane (Bromoform)	CHBr3	Thick	7 - 200	10	20 sec	2	-	Trichloroethan e-5	D5086834	10
Tribromomethane, 1,1,1	CHBr3	Thick	5 - 1500	?	20 sec	2	-	Trichloroethan e-5	D5086834	10
Trichloroethane, 1,1,1	CHBr3	Thick	5 - 300	10	20 sec	2	-	Trichloroethan e-5	D5086834	10
Trichloroethane, 1,1,1	CHBr3	Thick	n/a	n/a	n/a	2	-	Trichloroethan e-5-ETW	D5093834	10
Trichloroethane, 1,1,2	CHBr3	Thick	10 - 170	10	20 sec	2	-	Trichloroethan e-5	D5086834	10
Trichloroethane, 1,1,2	CHBr3	Thick	n/a	n/a	n/a	2	-	1,1,2- Trichloroethan e-ETW	D5094834	10
Trichloroethylene	C2HCl3	Thick	Warning Marks	n/a	n/a	2	-	CKW-A	D5085771	10
Trichloroethylene	C2HCl3	Thick	20 - 250	1	25 sec	2	-	Tri-5	D5085842	10
Trichloroethylene	C2HCl3	Thick	5 - 60	3	25 sec	2	-	Tri-5	D5085842	10
Trichloroethylene	C2HCl3	Thick	20 - 250	1	25 sec	2	-	Tri-5	D5085942	200
Trichloroethylene	C2HCl3	Thick	5 - 60	3	25 sec	2	-	Tri-5	D5085942	200

Trichloroethylene	C2HCl3	Thick	n/a	n/a	n/a	2	-	Tri-5-ETW	D5092842	10
Trichloroethylene	C2HCl3	Thin	20 - 250	1	1,5 min	1	-	SCE-4L	UNI-CF000143	5
Trichloroethylene	C2HCl3	Thin	0.05 - 1 %	1	3 min	1	-	SCE-6	UNI-CF000144	5
Trichloroethylene	C2HCl3	Thin	0.5 - 8	2	3 min	1	-	SCE-2M	UNI-CF000671	5
Trichloroethylene	C2HCl3	Thin	1 - 16	1	3 min	1	-	SCE-2M	UNI-CF000671	10
Trichloroethylene	C2HCl3	Thin	2 - 32	½	3 min	1	-	SCE-2M	UNI-CF000671	10
Trichloromethane (Chloroform)	CHCl3	Thick	8 - 100	10	20 sec	2	-	Trichloroethan e-5	D5086834	10
Trichloropropane, 1,2,3		Thick	10 - 1200	10	20 sec	2	-	Trichloroethan e-5	D5086834	10
Trimethyl Pentane, 2,2,4		Thick	100 - 3000	2	20 sec	2	-	Hexane-20	D5086832	10
UDMH		Thin	0.1 - 2	5	1 min	2	H	SHY-1L	UNI-CF000070	10
Vinyl Chloride	C2H3Cl	Thick	Warning Marks	n/a	n/a	2	-	CKW-A	D5085771	10
Vinyl Chloride	C2H3Cl	Thick	5 - 70	2	25 sec	2	-	VC-1	D5085837	10
Vinyl Chloride	C2H3Cl	Thick	1 - 15	10	25 sec	2	-	VC-1	D5085837	10
Vinyl Chloride	C2H3Cl	Thick	n/a	n/a	n/a	2	-	VC-1-ETW	D5092837	10
Vinyl Chloride	C2H3Cl	Thin	1 - 20	2	2 min	1	-	SVC-2M	UNI-CF000650	5
Vinyl Chloride	C2H3Cl	Thin	0.05-1%	2	1 min	1	-	SVC-6	UNI-CF000691	5
VX	C11H26NO 2PS	Thick	>0,0005 mg/m3	60	10 sec	5	-	CWA Sarin Tubes	7Sol-002	10
VX	C11H26NO 2PS	Thick	>0,0005 mg/m3	50	10 sec	2	-	CWA-GB.GX.VX- Tubes	7Sol-002	10
VX	C11H26NO 2PS	Thick	>0,5 mg/m3	5	10 sec	2	-	CWA-GB.GX.VX- Tubes	7Sol-002	10
VX	C11H26NO 2PS	Thick	0,0005 - 10 mg/m3	50	10 sec	1	-	CWA- GB.GD.GP.GA.G F.VX-Tubes	7Sol-008	10
VX	C11H26NO 2PS	Thick	0,005 - 100 mg/m3	5	10 sec	1	-	CWA- GB.GD.GP.GA.G F.VX-Tubes	7Sol-008	10
VX	C11H26NO 2PS	Thick	>0,01 mg/m3	10	10 sec	1	-	Phosph. Acid Esters-CWA	D5086880	10

Water Vapour	H2O	Thick	5 - 70 mg/m3	10 liter	10 min	2	-	H2O-HP (mg/m3)	D5085849	10
Water Vapour	H2O	Thick	20 - 160 mg/m3	5 liter	5 min	2	-	H2O-HP (mg/m3)	D5085849	10
Water Vapour	H2O	Thick	10 - 80	10 liter	10 min	2	-	H2O-HP (ppm)	D5085851	10
Water Vapour	H2O	Thick	30 - 200	5 liter	5 min	2	-	H2O-HP (ppm)	D5085851	10
Water Vapour	H2O	Thick	10 - 100 %	7	20	2	-	H2O (r.h%)	D5086849	10
Water Vapour	H2O	Thick	100 - 1000	n/a	n/a	2	-	H2O-MP	D5086851	10
Water Vapour	H2O	Thin	0.1 - 2 mg/l	1	1 min	1	-	SWA-5	UNI-CF000145	10
Water Vapour	H2O	Thin	1 - 18 mg/l	1	1 min	1	-	SWA-7M	UNI-CF000146	10
Water Vapour	H2O	Thin	1 - 30 mg/l	1	1 min	1	-	SWA-7	UNI-CF000147	10
Water Vapour	H2O	Thin	0.016 - 0.08 mg/l	4	1,5 min	1	-	SWA-4L	UNI-CF000712	10
Water Vapour	H2O	Thin	0.032 - 0.16 mg/l	2	1,5 min	1	-	SWA-4L	UNI-CF000712	10
Water Vapour	H2O	Thin	0.064 - 0.32 mg/l	1	1,5 min	1	-	SWA-4L	UNI-CF000712	10
Water Vapour	H2O	Thin	0.048 - 0.32 mg/l	2	1,5 min	1	-	SWA-4	UNI-CF000715	10
Water Vapour	H2O	Thin	0.096 - 0.64 mg/l	1	1,5 min	1	-	SWA-4	UNI-CF000715	10
Water Vapour	H2O	Thin	0.192 - 1.28 mg/l	½	1,5 min	1	-	SWA-4	UNI-CF000715	10
Xylene (Methyl Toluene)	C8H10	Thick	50 - 630	10	15 sec	2	-	Aromatic-HC	D5086811	10
Xylene (Methyl Toluene)	C24H30	Thick	50 - 630	10	15 sec	2	-	Aromatic-HC	D5086811	10
Xylene (Methyl Toluene)	C8H10	Thin	25 - 1000	1	2,5 min	2	-	SXY-4	UNI-CF000148	10
Xylene (Methyl Toluene)	C8H10	Thin	50 - 2000	½	2,5 min	2	-	SXY-4	UNI-CF000148	10
Xylene, 1,2 (Methyl Toluene)	C24H30	Thick	Warning Marks	n/a	n/a	2	-	Aromatic-A	D5085770	10
Xylene, 1,2 (Methyl Toluene)	C24H30	Thick	5 - 2500	10	40 sec	3	-	Tol-5	D5085828	10

Xylene, 1,3 (Methyl Toluene)	C24H30	Thick	Warning Marks	n/a	n/a	2	-	Aromatic-A	D5085770	10
Xylene, 1,3 (Methyl Toluene)	C24H30	Thick	50 - 630	10	15 sec	2	-	Aromatic HC	D5086811	10
Xylene, 1,3 (Methyl Toluene)	C24H30	Thick	5 - 2500	10	40 sec	3	-	Tol-5	D5085828	10
Xylene, 1,4 (Methyl Toluene)	C24H30	Thick	Warning Marks	n/a	n/a	2	-	Aromatic-A	D5085770	10
Xylene, 1,4 (Methyl Toluene)	C24H30	Thick	5 - 1200	10	40 sec	3	-	Tol-5	D5085828	10
Xylene, 1,4 (Methyl Toluene)	C24H30	Thick	50 - 630	10	15 sec	2	-	Aromatic-HC	D5086811	10

Important information

Tubes Storage: For a longer period of time, tubes should always be stored at a temperature below 10 °C temperature, and some tubes even

Warranty: The ASP-40 Air Sampling Pump will be provided with a 5 year warranty.

Payment: Payment shall be done within 30 days netto from day of order.

General Conditions of Delivery: The General Conditions of Delivery issued by FHI, Federation of Technology Branches, will be applied on all our transactions, as filed with the Clerk of the District Court of Utrecht and at the Chamber of Commerce in Amersfoort, under the number of 40507574. Exceptions are only valid after written confirmation.

Batch No.: From each dispatched package of gas detection tubes, we will hold one package per batch number as a reference at the location where the package has been dispatched from. To be sure every product we dispatch works properly, we test every batch number before dispatching.

Discarding Tubes: To protect and keep our environment clean of chemical waste, gas detection tubes should always be discarded with the Small Chemical Waste according to local regulations.

